

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020924\
 Data File : VY017341.D
 Acq On : 09 Feb 2024 15:56
 Operator : SY/MD
 Sample : P1423-01
 Misc : 5.14g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-3

Quant Time: Feb 09 22:04:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	58531	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	111163	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	98194	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	35626	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	28254	54.035	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.080%
35) Dibromofluoromethane	7.715	113	29243	48.721	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.440%
50) Toluene-d8	10.178	98	101635	46.508	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.020%
62) 4-Bromofluorobenzene	12.483	95	28082	39.228	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	78.460%
Target Compounds						
20) Methylene Chloride	4.704	84	6355	7.090	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020924\
Data File : VY017341.D
Acq On : 09 Feb 2024 15:56
Operator : SY/MD
Sample : P1423-01
Misc : 5.14g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-3

Quant Time: Feb 09 22:04:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

