



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021022\
Data File : VY007461.D
Acq On : 10 Feb 2022 09:08
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
Title : SW846 8260
Last Update : Thu Feb 03 04:36:48 2022



AutoFind: Scans 1749, 1750, 1751; Background Corrected with Scan 1741

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	18.3	11631	PASS
75	95	30	60	54.6	34675	PASS
95	95	100	100	100.0	63512	PASS
96	95	5	9	6.3	4021	PASS
173	174	0.00	2	1.3	635	PASS
174	95	50	100	77.1	48987	PASS
175	174	5	9	7.2	3522	PASS
176	174	95	101	99.3	48635	PASS
177	176	5	9	6.5	3165	PASS