

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021025\
 Data File : VY021154.D
 Acq On : 10 Feb 2025 13:51
 Operator : SY/MD
 Sample : Q1340-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RV-SOIL

Quant Time: Feb 11 00:30:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	165735	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.621	114	308833	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	288207	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	115466	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.066	65	96741	56.971	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 113.940%	
35) Dibromofluoromethane	7.640	113	101223	51.143	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.280%	
50) Toluene-d8	10.109	98	382030	50.701	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.400%	
62) 4-Bromofluorobenzene	12.407	95	122171	49.630	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 99.260%	
Target Compounds						
16) Acetone	3.878	43	5186	19.584	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021025\
Data File : VY021154.D
Acq On : 10 Feb 2025 13:51
Operator : SY/MD
Sample : Q1340-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RV-SOIL

Quant Time: Feb 11 00:30:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 03 13:08:38 2025
Response via : Initial Calibration

