

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021125\
 Data File : VY021168.D
 Acq On : 11 Feb 2025 12:21
 Operator : SY/MD
 Sample : Q1341-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HAR-ROLLOFFS

Quant Time: Feb 12 01:02:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	165178	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	308086	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	282384	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	105411	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	96494	57.017	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.040%
35) Dibromofluoromethane	7.646	113	99236	50.261	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.520%
50) Toluene-d8	10.115	98	376382	50.072	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.140%
62) 4-Bromofluorobenzene	12.414	95	115984	47.231	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	94.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021125\
Data File : VY021168.D
Acq On : 11 Feb 2025 12:21
Operator : SY/MD
Sample : Q1341-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HAR-ROLLOFFS

Quant Time: Feb 12 01:02:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 03 13:08:38 2025
Response via : Initial Calibration

