

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021125\
 Data File : VY021171.D
 Acq On : 11 Feb 2025 15:38
 Operator : SY/MD
 Sample : Q1345-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-01-021025

Quant Time: Feb 12 01:03:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	20897	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	31399	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	24093	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.359	152	7136	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	9451	44.142	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 88.280%			
35) Dibromofluoromethane	7.640	113	10328	51.325	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 102.660%			
50) Toluene-d8	10.115	98	36103	47.127	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 94.260%			
62) 4-Bromofluorobenzene	12.414	95	8303	33.176	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 66.360%			
Target Compounds						
80) 1,3,5-Trimethylbenzene	12.749	105	1128	2.496	ug/l	92
84) 1,2,4-Trimethylbenzene	13.054	105	1283	2.903	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021125\
Data File : VY021171.D
Acq On : 11 Feb 2025 15:38
Operator : SY/MD
Sample : Q1345-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-01-021025

Quant Time: Feb 12 01:03:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 03 13:08:38 2025
Response via : Initial Calibration

