

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021422\
 Data File : VY007493.D
 Acq On : 14 Feb 2022 13:39
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Feb 15 03:46:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y021422S.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 15 03:43:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	109461	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	204483	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	193168	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	80371	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	63631	53.008	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.020%
35) Dibromofluoromethane	7.722	113	64812	50.266	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.540%
50) Toluene-d8	10.185	98	246373	50.267	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.540%
62) 4-Bromofluorobenzene	12.483	95	86072	51.477	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021422\
Data File : VY007493.D
Acq On : 14 Feb 2022 13:39
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Feb 15 03:46:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y021422S.M
Quant Title : SW846 8260
QLast Update : Tue Feb 15 03:43:03 2022
Response via : Initial Calibration

