

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021522\
 Data File : VY007499.D
 Acq On : 15 Feb 2022 12:17
 Operator : SY/MD
 Sample : VY0215SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0215SBL01

Quant Time: Feb 16 01:38:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y021422S.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 15 09:34:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	107271	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	197412	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	186897	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	79047	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	64156	54.171	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.340%
35) Dibromofluoromethane	7.728	113	63744	51.203	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.400%
50) Toluene-d8	10.185	98	237159	50.156	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.320%
62) 4-Bromofluorobenzene	12.483	95	82161	51.245	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021522\
Data File : VY007499.D
Acq On : 15 Feb 2022 12:17
Operator : SY/MD
Sample : VY0215SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0215SBL01

Quant Time: Feb 16 01:38:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y021422S.M
Quant Title : SW846 8260
QLast Update : Tue Feb 15 09:34:50 2022
Response via : Initial Calibration

