

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021723\
 Data File : VY012639.D
 Acq On : 17 Feb 2023 10:06
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 17 11:44:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 01:09:03 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	84	0.00
2 T	Dichlorodifluoromethane	0.403	0.316	21.6	78	0.00
3 P	Chloromethane	0.390	0.343	12.1	83	0.00
4 C	Vinyl Chloride	0.485	0.447	7.8#	85	0.00
5 T	Bromomethane	0.400	0.356	11.0	87	0.00
6 T	Chloroethane	0.338	0.328	3.0	89	0.00
7 T	Trichlorofluoromethane	0.998	0.999	-0.1	90	0.00
8 T	Diethyl Ether	0.296	0.297	-0.3	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.504	0.504	0.0	88	-0.01
10 T	Methyl Iodide	0.653	0.630	3.5	79	0.00
11 T	Tert butyl alcohol	0.072	0.057	20.8	94	0.00
12 CM	1,1-Dichloroethene	0.498	0.483	3.0#	87	0.00
13 T	Acrolein	0.075	0.070	6.7	89	0.00
14 T	Allyl chloride	0.607	0.576	5.1	84	0.00
15 T	Acrylonitrile	0.136	0.139	-2.2	93	0.00
16 T	Acetone	0.153	0.165	-7.8	93	0.00
17 T	Carbon Disulfide	1.291	1.045	19.1	76	0.00
18 T	Methyl Acetate	0.368	0.337	8.4	91	0.00
19 T	Methyl tert-butyl Ether	1.457	1.580	-8.4	96	0.00
20 T	Methylene Chloride	0.677	0.575	15.1	85	0.00
21 T	trans-1,2-Dichloroethene	0.543	0.535	1.5	87	0.00
22 T	Diisopropyl ether	1.234	1.272	-3.1	88	-0.01
23 T	Vinyl Acetate	0.744	0.773	-3.9	89	0.00
24 P	1,1-Dichloroethane	0.882	0.886	-0.5	88	0.00
25 T	2-Butanone	0.170	0.170	0.0	93	0.00
26 T	2,2-Dichloropropane	0.969	1.016	-4.9	92	0.00
27 T	cis-1,2-Dichloroethene	0.630	0.658	-4.4	91	0.00
28 T	Bromochloromethane	0.321	0.340	-5.9	87	0.00
29 T	Tetrahydrofuran	0.092	0.088	4.3	89	0.00
30 C	Chloroform	1.064	1.106	-3.9#	91	0.00
31 T	Cyclohexane	0.735	0.645	12.2	80	0.00
32 T	1,1,1-Trichloroethane	1.032	1.091	-5.7	91	0.00
33 S	1,2-Dichloroethane-d4	0.636	0.653	-2.7	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.364	0.404	-11.0	90	0.00
36 T	1,1-Dichloropropene	0.493	0.497	-0.8	86	0.00
37 T	Ethyl Acetate	0.222	0.211	5.0	86	0.00
38 T	Carbon Tetrachloride	0.662	0.715	-8.0	93	0.00
39 T	Methylcyclohexane	0.593	0.585	1.3	83	0.00
40 TM	Benzene	1.456	1.496	-2.7	88	0.00
41 T	Methacrylonitrile	0.122	0.135	-10.7	87	0.00
42 TM	1,2-Dichloroethane	0.472	0.514	-8.9	94	0.00
43 T	Isopropyl Acetate	0.412	0.433	-5.1	93	0.00
44 TM	Trichloroethene	0.443	0.466	-5.2	91	0.00
45 C	1,2-Dichloropropane	0.316	0.326	-3.2#	87	0.00
46 T	Dibromomethane	0.209	0.220	-5.3	92	0.00
47 T	Bromodichloromethane	0.542	0.587	-8.3	92	0.00
48 T	Methyl methacrylate	0.188	0.197	-4.8	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021723\
 Data File : VY012639.D
 Acq On : 17 Feb 2023 10:06
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 17 11:44:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 01:09:03 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	102	0.00
50 S	Toluene-d8	1.516	1.609	-6.1	86	0.00
51 T	4-Methyl-2-Pentanone	0.222	0.233	-5.0	93	0.00
52 CM	Toluene	0.987	1.031	-4.5#	88	0.00
53 T	t-1,3-Dichloropropene	0.546	0.596	-9.2	92	0.00
54 T	cis-1,3-Dichloropropene	0.591	0.627	-6.1	90	0.00
55 T	1,1,2-Trichloroethane	0.305	0.331	-8.5	94	0.00
56 T	Ethyl methacrylate	0.384	0.420	-9.4	93	0.00
57 T	1,3-Dichloropropane	0.493	0.525	-6.5	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.172	0.190	-10.5	94	0.00
59 T	2-Hexanone	0.168	0.177	-5.4	92	0.00
60 T	Dibromochloromethane	0.407	0.460	-13.0	96	0.00
61 T	1,2-Dibromoethane	0.286	0.305	-6.6	93	0.00
62 S	4-Bromofluorobenzene	0.529	0.557	-5.3	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	82	0.00
64 T	Tetrachloroethene	0.522	0.607	-16.3	94	0.00
65 PM	Chlorobenzene	1.186	1.304	-9.9	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.465	0.527	-13.3	96	0.00
67 C	Ethyl Benzene	2.082	2.254	-8.3#	90	0.00
68 T	m/p-Xylenes	0.831	0.908	-9.3	90	0.00
69 T	o-Xylene	0.791	0.894	-13.0	92	0.00
70 T	Styrene	1.342	1.515	-12.9	93	0.00
71 P	Bromoform	0.294	0.332	-12.9	96	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	4.045	4.215	-4.2	91	0.00
74 T	N-amyl acetate	0.732	0.747	-2.0	93	0.00
75 P	1,1,2,2-Tetrachloroethane	0.654	0.638	2.4	93	0.00
76 T	1,2,3-Trichloropropane	0.559	0.572	-2.3	98	0.00
77 T	Bromobenzene	0.980	1.031	-5.2	96	0.00
78 T	n-propylbenzene	4.656	4.802	-3.1	91	0.00
79 T	2-Chlorotoluene	2.643	2.713	-2.6	92	0.00
80 T	1,3,5-Trimethylbenzene	3.482	3.644	-4.7	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.248	0.248	0.0	93	0.00
82 T	4-Chlorotoluene	2.773	2.867	-3.4	92	0.00
83 T	tert-Butylbenzene	3.054	3.258	-6.7	94	0.00
84 T	1,2,4-Trimethylbenzene	3.426	3.603	-5.2	91	0.00
85 T	sec-Butylbenzene	4.369	4.597	-5.2	92	0.00
86 T	p-Isopropyltoluene	3.790	4.081	-7.7	94	0.00
87 T	1,3-Dichlorobenzene	1.951	2.067	-5.9	95	0.00
88 T	1,4-Dichlorobenzene	1.954	2.044	-4.6	95	0.00
89 T	n-Butylbenzene	3.239	3.434	-6.0	91	0.00
90 T	Hexachloroethane	0.696	0.708	-1.7	91	0.00
91 T	1,2-Dichlorobenzene	1.755	1.845	-5.1	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.130	0.137	-5.4	103	0.00
93 T	1,2,4-Trichlorobenzene	1.087	1.213	-11.6	100	0.00
94 T	Hexachlorobutadiene	0.676	0.735	-8.7	98	0.00
95 T	Naphthalene	2.087	2.370	-13.6	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021723\
Data File : VY012639.D
Acq On : 17 Feb 2023 10:06
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Feb 17 11:44:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 01:09:03 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.941	1.036	-10.1	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6