



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021723\
Data File : VY012638.D
Acq On : 17 Feb 2023 09:00
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
Title : SW846 8260
Last Update : Fri Feb 17 18:08:19 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	12779	PASS
75	95	30	60	52.0	39923	PASS
95	95	100	100	100.0	76747	PASS
96	95	5	9	6.5	4955	PASS
173	174	0.00	2	1.1	838	PASS
174	95	50	100	97.9	75133	PASS
175	174	5	9	7.5	5640	PASS
176	174	95	101	97.0	72877	PASS
177	176	5	9	6.6	4806	PASS