

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021822\
 Data File : VY007547.D
 Acq On : 18 Feb 2022 16:04
 Operator : SY/MD
 Sample : N1580-08
 Misc : 5.88g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 D-8

Quant Time: Feb 19 04:33:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y021422S.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 15 09:34:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	73575	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	125059	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	122187	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	53923	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	44041	54.217	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.440%	
35) Dibromofluoromethane	7.722	113	41937	53.175	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.360%	
50) Toluene-d8	10.178	98	152574	50.936	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 101.880%	
62) 4-Bromofluorobenzene	12.477	95	56193	55.326	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 110.660%	
Target Compounds						Qvalue
39) Methylcyclohexane	9.179	83	9348	5.635	ug/l	90
80) 1,3,5-Trimethylbenzene	12.812	105	5655	1.429	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021822\
Data File : VY007547.D
Acq On : 18 Feb 2022 16:04
Operator : SY/MD
Sample : N1580-08
Misc : 5.88g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
D-8

Quant Time: Feb 19 04:33:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y021422S.M
Quant Title : SW846 8260
QLast Update : Tue Feb 15 09:34:50 2022
Response via : Initial Calibration

