

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022025\
 Data File : VY021260.D
 Acq On : 20 Feb 2025 19:26
 Operator : SY/MD
 Sample : Q1383-19
 Misc : 6.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-PCS-TC-20-021725

Quant Time: Feb 21 02:45:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.720	168	185650	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	358148	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	329429	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.359	152	128849	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	112982	59.397	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.800%
35) Dibromofluoromethane	7.646	113	117691	51.276	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.560%
50) Toluene-d8	10.115	98	438898	50.228	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.460%
62) 4-Bromofluorobenzene	12.414	95	138930	48.667	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	97.340%
Target Compounds						
16) Acetone	3.891	43	8394	28.298	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022025\
Data File : VY021260.D
Acq On : 20 Feb 2025 19:26
Operator : SY/MD
Sample : Q1383-19
Misc : 6.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-PCS-TC-20-021725

Quant Time: Feb 21 02:45:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 03 13:08:38 2025
Response via : Initial Calibration

