

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022025\
 Data File : VY021264.D
 Acq On : 20 Feb 2025 21:00
 Operator : SY/MD
 Sample : Q1382-11
 Misc : 5.75g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-PCS-TC-06-021725

Quant Time: Feb 21 02:47:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	168061	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	325406	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	303945	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	119761	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	105357	61.186	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	122.380%	
35) Dibromofluoromethane	7.646	113	106723	51.176	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	102.360%	
50) Toluene-d8	10.115	98	402696	50.722	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	101.440%	
62) 4-Bromofluorobenzene	12.414	95	127552	49.177	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	98.360%	
Target Compounds						
					Qvalue	
16) Acetone	3.885	43	21390	79.656	ug/l #	87
18) Methyl Acetate	4.403	43	1134	1.203	ug/l #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022025\
Data File : VY021264.D
Acq On : 20 Feb 2025 21:00
Operator : SY/MD
Sample : Q1382-11
Misc : 5.75g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-PCS-TC-06-021725

Quant Time: Feb 21 02:47:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 03 13:08:38 2025
Response via : Initial Calibration

