

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022025\
 Data File : VY021267.D
 Acq On : 20 Feb 2025 22:10
 Operator : SY/MD
 Sample : Q1382-17
 Misc : 4.25g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-PCS-TC-09-021725

Quant Time: Feb 21 02:48:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.720	168	179849	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.628	114	348924	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	316031	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.359	152	118848	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	105356	57.175	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.340%
35) Dibromofluoromethane	7.646	113	112496	50.308	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.620%
50) Toluene-d8	10.115	98	427049	50.164	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.320%
62) 4-Bromofluorobenzene	12.414	95	129419	46.534	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	93.060%
Target Compounds						
						Qvalue
16) Acetone	3.891	43	20515	71.390	ug/l	100
18) Methyl Acetate	4.391	43	1206	1.196	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022025\
Data File : VY021267.D
Acq On : 20 Feb 2025 22:10
Operator : SY/MD
Sample : Q1382-17
Misc : 4.25g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-PCS-TC-09-021725

Quant Time: Feb 21 02:48:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 03 13:08:38 2025
Response via : Initial Calibration

