



Bromomethane

Response Ratio



$R = -4.819e-002 A^2 + 4.409e-001 A - 1.288e-004$
 Coef of Det (r^2) = 0.996064 Curve Fit: Quadratic
 Method Name: Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022122S.M
 Calibration Table Last Updated: Mon Feb 21 15:25:36 2022


Methyl Acetate



Response = $2.315 \times 10^{-1} \times \text{Amt} + 5.631 \times 10^{-3}$
 Coef of Det (r^2) = 0.993274 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022122S.M
 Calibration Table Last Updated: Mon Feb 21 15:25:36 2022

Methylene Chloride

Response Ratio



$$\text{Response} = 4.773\text{e-}001 * \text{Amt} + 3.888\text{e-}002$$


Coef of Det (r^2) = 0.998835 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022122S.M

Calibration Table Last Updated: Mon Feb 21 15:25:36 2022

Diisopropyl ether

Response Ratio



$$\text{Response} = 9.327\text{e-}001 * \text{Amt} + 6.438\text{e-}002$$


Coef of Det (r^2) = 0.996331 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022122S.M

Calibration Table Last Updated: Mon Feb 21 15:25:36 2022

Cyclohexane

Response Ratio



$$\text{Response} = 5.588\text{e-}001 * \text{Amt} + 6.821\text{e-}002$$

Coef of Det (r^2) = 0.998155 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022122S.M

Calibration Table Last Updated: Mon Feb 21 15:25:36 2022