



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022122\
 Data File : VY007560.D
 Acq On : 21 Feb 2022 08:51
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022122S.M
 Title : SW846 8260
 Last Update : Mon Feb 21 15:25:36 2022



AutoFind: Scans 1749, 1750, 1751; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.9	7228	PASS
75	95	30	60	52.8	22549	PASS
95	95	100	100	100.0	42688	PASS
96	95	5	9	6.4	2728	PASS
173	174	0.00	2	1.6	590	PASS
174	95	50	100	87.4	37315	PASS
175	174	5	9	7.1	2641	PASS
176	174	95	101	95.5	35648	PASS
177	176	5	9	6.2	2196	PASS