

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022125\
 Data File : VY021286.D
 Acq On : 21 Feb 2025 15:24
 Operator : SY/MD
 Sample : Q1383-15
 Misc : 5.74g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-PCS-TC-18-021725

Quant Time: Feb 21 23:08:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	168253	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	327688	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	315012	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	126536	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	112157	65.061	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 130.120%	
35) Dibromofluoromethane	7.640	113	110308	52.527	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.060%	
50) Toluene-d8	10.115	98	407398	50.957	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.920%	
62) 4-Bromofluorobenzene	12.414	95	132835	50.857	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 101.720%	
Target Compounds						
16) Acetone	3.873	43	29983	111.529	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022125\
Data File : VY021286.D
Acq On : 21 Feb 2025 15:24
Operator : SY/MD
Sample : Q1383-15
Misc : 5.74g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-PCS-TC-18-021725

Quant Time: Feb 21 23:08:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 03 13:08:38 2025
Response via : Initial Calibration

