

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022222\
 Data File : VY007580.D
 Acq On : 22 Feb 2022 14:08
 Operator : SY/MD
 Sample : N1600-07RE
 Misc : 5.71g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 C0AA8RE

Quant Time: Feb 22 14:37:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 15:25:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	9203	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	13870	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	13544	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	6301	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	4511	47.327	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.660%
35) Dibromofluoromethane	7.716	113	4139	47.062	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.120%
50) Toluene-d8	10.179	98	16906	55.441	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.880%
62) 4-Bromofluorobenzene	12.483	95	5894	53.381	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022222\
Data File : VY007580.D
Acq On : 22 Feb 2022 14:08
Operator : SY/MD
Sample : N1600-07RE
Misc : 5.71g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
C0AA8RE

Quant Time: Feb 22 14:37:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022122S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 15:25:36 2022
Response via : Initial Calibration

