

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022222\
 Data File : VY007585.D
 Acq On : 22 Feb 2022 16:07
 Operator : SY/MD
 Sample : N1594-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RO-CONCRETE-COMP

Quant Time: Feb 22 17:03:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 15:25:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	71501	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	117388	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	122399	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	61324	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	42109	56.862	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.720%
35) Dibromofluoromethane	7.716	113	30541	41.031	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	82.060%
50) Toluene-d8	10.179	98	144716	56.073	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.140%
62) 4-Bromofluorobenzene	12.477	95	58309	62.397	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	124.800%
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.117	105	10884	2.721	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022222\
Data File : VY007585.D
Acq On : 22 Feb 2022 16:07
Operator : SY/MD
Sample : N1594-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RO-CONCRETE-COMP

Quant Time: Feb 22 17:03:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022122S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 15:25:36 2022
Response via : Initial Calibration

