

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022222\
 Data File : VY007587.D
 Acq On : 22 Feb 2022 16:54
 Operator : SY/MD
 Sample : N1635-01
 Misc : 4.53g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RO-124055

Quant Time: Feb 23 03:32:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 15:25:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	57246	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	92276	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	89538	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	45079	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	34508	58.202	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.400%
35) Dibromofluoromethane	7.710	113	31522	53.874	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.740%
50) Toluene-d8	10.179	98	109424	53.937	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.880%
62) 4-Bromofluorobenzene	12.477	95	41487	56.478	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.960%
Target Compounds						
						Qvalue
74) N-amyl acetate	12.148	43	2498	3.610	ug/l	# 47
84) 1,2,4-Trimethylbenzene	13.111	105	4367	1.485	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022222\
Data File : VY007587.D
Acq On : 22 Feb 2022 16:54
Operator : SY/MD
Sample : N1635-01
Misc : 4.53g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RO-124055

Quant Time: Feb 23 03:32:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022122S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 15:25:36 2022
Response via : Initial Calibration

