

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022222\
 Data File : VY007588.D
 Acq On : 22 Feb 2022 17:41
 Operator : SY/MD
 Sample : N1635-03
 Misc : 6.53g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RO-124042

Quant Time: Feb 23 03:32:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 15:25:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	60667	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	97968	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	96841	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	50541	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	34272	54.544	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.080%
35) Dibromofluoromethane	7.716	113	33187	53.424	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.840%
50) Toluene-d8	10.179	98	114944	53.366	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.740%
62) 4-Bromofluorobenzene	12.477	95	46270	59.329	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	118.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022222\
Data File : VY007588.D
Acq On : 22 Feb 2022 17:41
Operator : SY/MD
Sample : N1635-03
Misc : 6.53g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RO-124042

Quant Time: Feb 23 03:32:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022122S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 15:25:36 2022
Response via : Initial Calibration

