

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022322\
 Data File : VY007611.D
 Acq On : 23 Feb 2022 20:26
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Feb 24 08:25:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022322S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 24 07:13:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	92673	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	161410	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	159717	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	76232	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	50518	49.002	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.000%
35) Dibromofluoromethane	7.716	113	50709	49.789	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.580%
50) Toluene-d8	10.179	98	184227	48.974	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.940%
62) 4-Bromofluorobenzene	12.477	95	67005	52.030	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.060%
Target Compounds						
					Qvalue	
31) Cyclohexane	7.795	56	1733	Below Cal	#	18
39) Methylcyclohexane	9.191	83	25	Below Cal	#	16

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022322\
Data File : VY007611.D
Acq On : 23 Feb 2022 20:26
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Feb 24 08:25:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022322S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 24 07:13:58 2022
Response via : Initial Calibration

