

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022323\
 Data File : VY012706.D
 Acq On : 23 Feb 2023 14:17
 Operator : KP/MD
 Sample : 01663-01
 Misc : 4.10g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-D2

Quant Time: Feb 23 23:54:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 22 01:21:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	143226	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	264719	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	262379	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	127679	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	90523	53.514	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 107.020%	
35) Dibromofluoromethane	7.716	113	86998	48.657	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 97.320%	
50) Toluene-d8	10.179	98	319448	47.689	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 95.380%	
62) 4-Bromofluorobenzene	12.483	95	113725	53.053	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 106.100%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022323\
Data File : VY012706.D
Acq On : 23 Feb 2023 14:17
Operator : KP/MD
Sample : 01663-01
Misc : 4.10g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-D2

Quant Time: Feb 23 23:54:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 22 01:21:05 2023
Response via : Initial Calibration

