

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022423\
 Data File : VY012724.D
 Acq On : 24 Feb 2023 15:55
 Operator : KP/MD
 Sample : 01611-02
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-26-SB02

Quant Time: Feb 25 00:50:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 22 01:21:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	238505	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	431086	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	408027	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	179710	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	151526	53.793	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.580%
35) Dibromofluoromethane	7.716	113	137508	47.226	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.460%
50) Toluene-d8	10.179	98	516414	47.341	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.680%
62) 4-Bromofluorobenzene	12.483	95	171064	49.005	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.000%
Target Compounds						
16) Acetone	3.948	43	13622	14.820	ug/l	97
17) Carbon Disulfide	4.198	76	15459	1.823	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022423\
Data File : VY012724.D
Acq On : 24 Feb 2023 15:55
Operator : KP/MD
Sample : 01611-02
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-26-SB02

Quant Time: Feb 25 00:50:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 22 01:21:05 2023
Response via : Initial Calibration

