

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022525\
 Data File : VY021317.D
 Acq On : 25 Feb 2025 17:31
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY022525

Quant Time: Feb 26 03:42:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022525S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 26 02:09:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	50.000	53.871	-7.7	109	0.00
3 P	Chloromethane	50.000	53.523	-7.0	114	0.00
4 C	Vinyl Chloride	50.000	55.184	-10.4#	111	0.00
5 T	Bromomethane	50.000	52.483	-5.0	114	0.00
6 T	Chloroethane	50.000	54.804	-9.6	110	0.00
7 T	Trichlorofluoromethane	50.000	53.384	-6.8	108	0.00
8 T	Diethyl Ether	50.000	54.849	-9.7	109	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	54.188	-8.4	111	0.00
10 T	Methyl Iodide	50.000	61.361	-22.7	112	0.00
11 T	Tert butyl alcohol	250.000	221.762	11.3	101	0.00
12 CM	1,1-Dichloroethene	50.000	54.104	-8.2#	111	0.00
13 T	Acrolein	250.000	51.554	79.4#	108	0.00
14 T	Allyl chloride	50.000	55.361	-10.7	112	0.00
15 T	Acrylonitrile	250.000	267.274	-6.9	106	0.00
16 T	Acetone	250.000	227.197	9.1	81	0.00
17 T	Carbon Disulfide	50.000	54.581	-9.2	110	0.00
18 T	Methyl Acetate	50.000	52.372	-4.7	106	0.00
19 T	Methyl tert-butyl Ether	50.000	54.311	-8.6	107	0.00
20 T	Methylene Chloride	50.000	51.062	-2.1	108	0.00
21 T	trans-1,2-Dichloroethene	50.000	55.044	-10.1	111	0.00
22 T	Diisopropyl ether	50.000	55.595	-11.2	110	0.00
23 T	Vinyl Acetate	250.000	275.197	-10.1	107	0.00
24 P	1,1-Dichloroethane	50.000	55.004	-10.0	111	0.00
25 T	2-Butanone	250.000	253.246	-1.3	95	0.00
26 T	2,2-Dichloropropane	50.000	53.893	-7.8	110	0.00
27 T	cis-1,2-Dichloroethene	50.000	56.305	-12.6	112	0.00
28 T	Bromochloromethane	50.000	47.452	5.1	101	0.00
29 T	Tetrahydrofuran	250.000	260.172	-4.1	102	0.00
30 C	Chloroform	50.000	55.023	-10.0#	111	0.00
31 T	Cyclohexane	50.000	52.180	-4.4	111	0.00
32 T	1,1,1-Trichloroethane	50.000	55.084	-10.2	112	0.00
33 S	1,2-Dichloroethane-d4	50.000	47.453	5.1	102	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	107	0.00
35 S	Dibromofluoromethane	50.000	49.037	1.9	106	0.00
36 T	1,1-Dichloropropene	50.000	54.593	-9.2	112	0.00
37 T	Ethyl Acetate	50.000	50.864	-1.7	104	0.00
38 T	Carbon Tetrachloride	50.000	54.157	-8.3	111	0.00
39 T	Methylcyclohexane	50.000	54.397	-8.8	110	0.00
40 TM	Benzene	50.000	54.583	-9.2	111	0.00
41 T	Methacrylonitrile	50.000	49.244	1.5	99	0.00
42 TM	1,2-Dichloroethane	50.000	53.746	-7.5	110	0.00
43 T	Isopropyl Acetate	50.000	52.380	-4.8	104	0.00
44 TM	Trichloroethene	50.000	54.930	-9.9	112	0.00
45 C	1,2-Dichloropropane	50.000	54.464	-8.9#	110	0.00
46 T	Dibromomethane	50.000	54.539	-9.1	110	0.00
47 T	Bromodichloromethane	50.000	54.489	-9.0	110	0.00
48 T	Methyl methacrylate	50.000	55.076	-10.2	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022525\
 Data File : VY021317.D
 Acq On : 25 Feb 2025 17:31
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICSVY022525

Quant Time: Feb 26 03:42:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022525S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 26 02:09:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1053.076	-5.3	107	0.00
50 S	Toluene-d8	50.000	48.694	2.6	106	0.00
51 T	4-Methyl-2-Pentanone	250.000	262.628	-5.1	102	0.00
52 CM	Toluene	50.000	55.674	-11.3#	112	0.00
53 T	t-1,3-Dichloropropene	50.000	55.699	-11.4	108	0.00
54 T	cis-1,3-Dichloropropene	50.000	55.451	-10.9	111	0.00
55 T	1,1,2-Trichloroethane	50.000	53.263	-6.5	107	0.00
56 T	Ethyl methacrylate	50.000	55.679	-11.4	105	0.00
57 T	1,3-Dichloropropane	50.000	54.438	-8.9	109	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	228.990	8.4	85	0.00
59 T	2-Hexanone	250.000	264.133	-5.7	99	0.00
60 T	Dibromochloromethane	50.000	54.329	-8.7	108	0.00
61 T	1,2-Dibromoethane	50.000	53.942	-7.9	108	0.00
62 S	4-Bromofluorobenzene	50.000	48.549	2.9	105	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	107	0.00
64 T	Tetrachloroethene	50.000	52.940	-5.9	110	0.00
65 PM	Chlorobenzene	50.000	54.410	-8.8	111	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	54.517	-9.0	111	0.00
67 C	Ethyl Benzene	50.000	55.992	-12.0#	112	0.00
68 T	m/p-Xylenes	100.000	111.274	-11.3	111	0.00
69 T	o-Xylene	50.000	56.108	-12.2	111	0.00
70 T	Styrene	50.000	56.307	-12.6	110	0.00
71 P	Bromoform	50.000	53.590	-7.2	105	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	104	0.00
73 T	Isopropylbenzene	50.000	56.974	-13.9	112	0.00
74 T	N-ethyl acetate	50.000	55.509	-11.0	104	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	53.254	-6.5	105	0.00
76 T	1,2,3-Trichloropropane	50.000	44.570	10.9	88	0.00
77 T	Bromobenzene	50.000	55.296	-10.6	110	0.00
78 T	n-propylbenzene	50.000	56.783	-13.6	111	0.00
79 T	2-Chlorotoluene	50.000	55.813	-11.6	112	0.00
80 T	1,3,5-Trimethylbenzene	50.000	56.549	-13.1	111	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	53.216	-6.4	102	0.00
82 T	4-Chlorotoluene	50.000	54.757	-9.5	110	0.00
83 T	tert-Butylbenzene	50.000	55.900	-11.8	112	0.00
84 T	1,2,4-Trimethylbenzene	50.000	56.621	-13.2	111	0.00
85 T	sec-Butylbenzene	50.000	55.655	-11.3	110	0.00
86 T	p-Isopropyltoluene	50.000	56.427	-12.9	111	0.00
87 T	1,3-Dichlorobenzene	50.000	54.800	-9.6	110	0.00
88 T	1,4-Dichlorobenzene	50.000	54.098	-8.2	110	0.00
89 T	n-Butylbenzene	50.000	56.856	-13.7	110	0.00
90 T	Hexachloroethane	50.000	55.035	-10.1	112	0.00
91 T	1,2-Dichlorobenzene	50.000	54.545	-9.1	109	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	52.137	-4.3	100	0.00
93 T	1,2,4-Trichlorobenzene	50.000	56.703	-13.4	112	0.00
94 T	Hexachlorobutadiene	50.000	53.541	-7.1	109	0.00
95 T	Naphthalene	50.000	57.589	-15.2	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022525\
Data File : VY021317.D
Acq On : 25 Feb 2025 17:31
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY022525

Quant Time: Feb 26 03:42:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022525S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 26 02:09:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	56.980	-14.0	112	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6