



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022822\
 Data File : VY007653.D
 Acq On : 28 Feb 2022 15:50
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822W.M
 Title : SW846 8260
 Last Update : Mon Feb 21 07:57:15 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.5	41549	PASS
75	95	30	60	55.6	107493	PASS
95	95	100	100	100.0	193344	PASS
96	95	5	9	6.6	12821	PASS
173	174	0.00	2	1.4	2530	PASS
174	95	50	100	94.3	182229	PASS
175	174	5	9	7.9	14316	PASS
176	174	95	101	95.8	174613	PASS
177	176	5	9	6.2	10859	PASS