

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022825\
 Data File : VY021387.D
 Acq On : 28 Feb 2025 17:52
 Operator : SY/MD
 Sample : Q1441-01
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HD-02-022625

Quant Time: Feb 28 22:23:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022525S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 26 02:09:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	33160	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	54415	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	46106	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	14621	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	23003	61.150	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	122.300%	
35) Dibromofluoromethane	7.634	113	21817	61.155	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	122.300%	
50) Toluene-d8	10.109	98	61983	44.956	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	89.920%	
62) 4-Bromofluorobenzene	12.408	95	16832	36.275	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	72.560%	
Target Compounds						
20) Methylene Chloride	4.623	84	3014	7.724	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022825\
Data File : VY021387.D
Acq On : 28 Feb 2025 17:52
Operator : SY/MD
Sample : Q1441-01
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HD-02-022625

Quant Time: Feb 28 22:23:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022525S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 26 02:09:13 2025
Response via : Initial Calibration

