

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030122\
 Data File : VY007668.D
 Acq On : 01 Mar 2022 15:14
 Operator : SY/MD
 Sample : N1717-01
 Misc : 5.56g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ETGI-359

Quant Time: Mar 02 07:51:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 01 09:08:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	287354	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	513656	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	490246	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	183116	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	175921	58.478	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.960%
35) Dibromofluoromethane	7.716	113	170203	51.829	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.660%
50) Toluene-d8	10.179	98	635930	52.606	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.220%
62) 4-Bromofluorobenzene	12.477	95	196220	50.356	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030122\
Data File : VY007668.D
Acq On : 01 Mar 2022 15:14
Operator : SY/MD
Sample : N1717-01
Misc : 5.56g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ETGI-359

Quant Time: Mar 02 07:51:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 01 09:08:03 2022
Response via : Initial Calibration

