

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030122\
 Data File : VY007673.D
 Acq On : 01 Mar 2022 17:23
 Operator : SY/MD
 Sample : N1721-09
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DUM-5

Quant Time: Mar 02 07:52:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 01 09:08:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	280040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	501634	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	503086	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	236811	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	178628	60.929	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.860%
35) Dibromofluoromethane	7.715	113	169912	52.980	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.960%
50) Toluene-d8	10.178	98	634306	53.730	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.460%
62) 4-Bromofluorobenzene	12.477	95	208474	54.783	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	109.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030122\
Data File : VY007673.D
Acq On : 01 Mar 2022 17:23
Operator : SY/MD
Sample : N1721-09
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DUM-5

Quant Time: Mar 02 07:52:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 01 09:08:03 2022
Response via : Initial Calibration

