

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030122\
 Data File : VY007676.D
 Acq On : 01 Mar 2022 18:57
 Operator : SY/MD
 Sample : N1722-07
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DUM-9

Quant Time: Mar 02 07:52:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 01 09:08:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	44427	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	70206	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	63560	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	28120	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	23516	50.560	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 101.120%	
35) Dibromofluoromethane	7.722	113	23945	53.348	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.700%	
50) Toluene-d8	10.179	98	81405	49.270	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 98.540%	
62) 4-Bromofluorobenzene	12.483	95	24747	46.465	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 92.940%	
Target Compounds						
16) Acetone	3.942	43	1327	14.025	ug/l	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030122\
Data File : VY007676.D
Acq On : 01 Mar 2022 18:57
Operator : SY/MD
Sample : N1722-07
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DUM-9

Quant Time: Mar 02 07:52:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 01 09:08:03 2022
Response via : Initial Calibration

