

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030322\
 Data File : VY007702.D
 Acq On : 03 Mar 2022 14:21
 Operator : SY/MD
 Sample : N1751-03
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DUM-12

Quant Time: Mar 04 00:41:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 01 09:08:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	261218	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	475191	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	481383	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	221601	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	169454	61.964	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 123.920%	
35) Dibromofluoromethane	7.722	113	160824	52.937	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.880%	
50) Toluene-d8	10.179	98	600995	53.741	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.480%	
62) 4-Bromofluorobenzene	12.483	95	197295	54.731	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 109.460%	
Target Compounds						
16) Acetone	3.948	43	9096	16.350	ug/l	Qvalue # 55

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030322\
Data File : VY007702.D
Acq On : 03 Mar 2022 14:21
Operator : SY/MD
Sample : N1751-03
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DUM-12

Quant Time: Mar 04 00:41:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 01 09:08:03 2022
Response via : Initial Calibration

