

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030525\
 Data File : VY021418.D
 Acq On : 05 Mar 2025 17:18
 Operator : SY/MD
 Sample : Q1475-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-04-02282025

Quant Time: Mar 05 18:13:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	268882	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	508298	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	449100	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	163914	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	171804	60.453	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.900%
35) Dibromofluoromethane	7.634	113	172397	51.599	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.200%
50) Toluene-d8	10.109	98	613557	48.500	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.000%
62) 4-Bromofluorobenzene	12.408	95	183683	42.685	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	85.380%
Target Compounds						
						Qvalue
16) Acetone	3.873	43	4503	7.265	ug/l	94
20) Methylene Chloride	4.622	84	6453	1.983	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030525\
Data File : VY021418.D
Acq On : 05 Mar 2025 17:18
Operator : SY/MD
Sample : Q1475-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-04-02282025

Quant Time: Mar 05 18:13:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration

