

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030623\
 Data File : VY012819.D
 Acq On : 06 Mar 2023 13:35
 Operator : KP/MD
 Sample : 01671-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 T5-1

Quant Time: Mar 07 04:23:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	1663	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	2434	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	1752	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	274	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	1155	69.952	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	139.900%
35) Dibromofluoromethane	7.710	113	839	57.117	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.240%
50) Toluene-d8	10.179	98	2779	51.750	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.500%
62) 4-Bromofluorobenzene	12.483	95	568	31.337	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	62.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030623\
Data File : VY012819.D
Acq On : 06 Mar 2023 13:35
Operator : KP/MD
Sample : 01671-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
T5-1

Quant Time: Mar 07 04:23:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

