

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030623\
 Data File : VY012827.D
 Acq On : 06 Mar 2023 16:36
 Operator : KP/MD
 Sample : 01814-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-206

Quant Time: Mar 07 04:25:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	5431	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	7936	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	7021	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	2365	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	4268	79.150	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 158.300%	
35) Dibromofluoromethane	7.716	113	3820	79.760	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 159.520%#	
50) Toluene-d8	10.179	98	9838	56.188	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 112.380%	
62) 4-Bromofluorobenzene	12.477	95	2592	43.860	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 87.720%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030623\
Data File : VY012827.D
Acq On : 06 Mar 2023 16:36
Operator : KP/MD
Sample : 01814-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-206

Quant Time: Mar 07 04:25:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

