

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030623\
 Data File : VY012812.D
 Acq On : 06 Mar 2023 10:21
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 06 14:13:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.493	0.476	3.4	96	0.00
3 P	Chloromethane	0.589	0.558	5.3	98	0.00
4 C	Vinyl Chloride	0.540	0.542	-0.4#	98	0.00
5 T	Bromomethane	0.328	0.331	-0.9	99	0.00
6 T	Chloroethane	0.324	0.330	-1.9	98	0.00
7 T	Trichlorofluoromethane	0.897	0.887	1.1	97	0.00
8 T	Diethyl Ether	0.314	0.296	5.7	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.540	0.531	1.7	96	0.00
10 T	Methyl Iodide	0.703	0.731	-4.0	93	0.00
11 T	Tert butyl alcohol	0.070	0.051	27.1#	86	0.00
12 CM	1,1-Dichloroethene	0.515	0.520	-1.0#	97	0.00
13 T	Acrolein	0.087	0.082	5.7	92	0.00
14 T	Allyl chloride	0.951	0.894	6.0	89	0.00
15 T	Acrylonitrile	0.166	0.153	7.8	88	0.00
16 T	Acetone	0.195	0.224	-14.9	101	0.00
17 T	Carbon Disulfide	1.753	1.695	3.3	93	0.00
18 T	Methyl Acetate	0.580	0.497	14.3	80	0.00
19 T	Methyl tert-butyl Ether	1.330	1.332	-0.2	92	0.00
20 T	Methylene Chloride	0.714	0.592	17.1	92	0.00
21 T	trans-1,2-Dichloroethene	0.560	0.573	-2.3	98	0.00
22 T	Diisopropyl ether	1.727	1.644	4.8	88	0.00
23 T	Vinyl Acetate	1.073	1.010	5.9	85	0.00
24 P	1,1-Dichloroethane	0.984	0.951	3.4	92	0.00
25 T	2-Butanone	0.241	0.239	0.8	90	0.00
26 T	2,2-Dichloropropane	0.874	0.851	2.6	95	0.00
27 T	cis-1,2-Dichloroethene	0.600	0.607	-1.2	94	0.00
28 T	Bromochloromethane	0.360	0.337	6.4	87	0.00
29 T	Tetrahydrofuran	0.146	0.126	13.7	79	0.00
30 C	Chloroform	0.972	0.972	0.0	95	0.00
31 T	Cyclohexane	0.951	0.872	8.3	90	0.00
32 T	1,1,1-Trichloroethane	0.879	0.867	1.4	94	0.00
33 S	1,2-Dichloroethane-d4	0.496	0.440	11.3	85	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.302	0.288	4.6	89	0.00
36 T	1,1-Dichloropropene	0.506	0.507	-0.2	94	0.00
37 T	Ethyl Acetate	0.319	0.285	10.7	86	0.00
38 T	Carbon Tetrachloride	0.522	0.537	-2.9	96	0.00
39 T	Methylcyclohexane	0.601	0.611	-1.7	92	0.00
40 TM	Benzene	1.430	1.445	-1.0	94	0.00
41 T	Methacrylonitrile	0.172	0.148	14.0	80	0.00
42 TM	1,2-Dichloroethane	0.430	0.417	3.0	90	0.00
43 T	Isopropyl Acetate	0.569	0.512	10.0	82	0.00
44 TM	Trichloroethene	0.377	0.387	-2.7	95	0.00
45 C	1,2-Dichloropropane	0.361	0.352	2.5#	92	0.00
46 T	Dibromomethane	0.206	0.207	-0.5	92	0.00
47 T	Bromodichloromethane	0.485	0.485	0.0	93	0.00
48 T	Methyl methacrylate	0.262	0.242	7.6	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030623\
 Data File : VY012812.D
 Acq On : 06 Mar 2023 10:21
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 06 14:13:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.002	33.3#	84	0.00
50 S	Toluene-d8	1.103	1.036	6.1	89	0.00
51 T	4-Methyl-2-Pentanone	0.308	0.281	8.8	81	0.00
52 CM	Toluene	0.863	0.901	-4.4#	96	0.00
53 T	t-1,3-Dichloropropene	0.509	0.510	-0.2	92	0.00
54 T	cis-1,3-Dichloropropene	0.576	0.581	-0.9	92	0.00
55 T	1,1,2-Trichloroethane	0.287	0.284	1.0	92	0.00
56 T	Ethyl methacrylate	0.382	0.376	1.6	86	0.00
57 T	1,3-Dichloropropane	0.489	0.482	1.4	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.201	0.198	1.5	87	0.00
59 T	2-Hexanone	0.224	0.224	0.0	87	0.00
60 T	Dibromochloromethane	0.349	0.351	-0.6	93	0.00
61 T	1,2-Dibromoethane	0.276	0.271	1.8	91	0.00
62 S	4-Bromofluorobenzene	0.372	0.353	5.1	89	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.364	0.370	-1.6	95	0.00
65 PM	Chlorobenzene	1.049	1.070	-2.0	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.388	0.399	-2.8	96	0.00
67 C	Ethyl Benzene	1.845	1.928	-4.5#	95	0.00
68 T	m/p-Xylenes	0.710	0.746	-5.1	95	0.00
69 T	o-Xylene	0.662	0.701	-5.9	97	0.00
70 T	Styrene	1.105	1.181	-6.9	95	0.00
71 P	Bromoform	0.253	0.252	0.4	93	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.544	3.728	-5.2	96	0.00
74 T	N-amyl acetate	1.115	1.021	8.4	82	0.00
75 P	1,1,2,2-Tetrachloroethane	0.797	0.769	3.5	90	0.00
76 T	1,2,3-Trichloropropane	0.597	0.587	1.7	92	0.00
77 T	Bromobenzene	0.861	0.884	-2.7	96	0.00
78 T	n-propylbenzene	4.369	4.600	-5.3	96	0.00
79 T	2-Chlorotoluene	2.428	2.506	-3.2	95	0.00
80 T	1,3,5-Trimethylbenzene	2.924	3.101	-6.1	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.288	0.276	4.2	88	0.00
82 T	4-Chlorotoluene	2.513	2.575	-2.5	94	0.00
83 T	tert-Butylbenzene	2.524	2.653	-5.1	95	0.00
84 T	1,2,4-Trimethylbenzene	2.860	3.052	-6.7	97	0.00
85 T	sec-Butylbenzene	3.851	4.027	-4.6	96	0.00
86 T	p-Isopropyltoluene	3.147	3.346	-6.3	97	0.00
87 T	1,3-Dichlorobenzene	1.660	1.725	-3.9	97	0.00
88 T	1,4-Dichlorobenzene	1.694	1.710	-0.9	95	0.00
89 T	n-Butylbenzene	3.013	3.148	-4.5	95	0.00
90 T	Hexachloroethane	0.655	0.675	-3.1	98	0.00
91 T	1,2-Dichlorobenzene	1.501	1.541	-2.7	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.129	0.119	7.8	83	0.00
93 T	1,2,4-Trichlorobenzene	0.924	0.933	-1.0	92	0.00
94 T	Hexachlorobutadiene	0.564	0.569	-0.9	97	0.00
95 T	Naphthalene	1.852	1.844	0.4	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030623\
Data File : VY012812.D
Acq On : 06 Mar 2023 10:21
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Mar 06 14:13:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.822	0.824	-0.2	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5