

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030722\
 Data File : VY007767.D
 Acq On : 07 Mar 2022 19:28
 Operator : SY/MD
 Sample : N1809-01
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-02-030422

Quant Time: Mar 08 03:04:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 01 09:08:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	177741	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	305620	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	276734	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	124254	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	119761	64.360	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 128.720%	
35) Dibromofluoromethane	7.716	113	102635	52.528	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.060%	
50) Toluene-d8	10.179	98	371183	51.607	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 103.220%	
62) 4-Bromofluorobenzene	12.477	95	114013	49.176	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 98.360%	
Target Compounds						
16) Acetone	3.954	43	4699	12.414	ug/l	Qvalue # 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030722\
Data File : VY007767.D
Acq On : 07 Mar 2022 19:28
Operator : SY/MD
Sample : N1809-01
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-02-030422

Quant Time: Mar 08 03:04:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 01 09:08:03 2022
Response via : Initial Calibration

