

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030725\
 Data File : VY021455.D
 Acq On : 07 Mar 2025 12:55
 Operator : SY/MD
 Sample : Q1463-03
 Misc : 5.45g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 T3

Quant Time: Mar 07 22:00:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.683	168	247407	50.000	ug/l	-0.03
34) 1,4-Difluorobenzene	8.603	114	315188	50.000	ug/l	-0.02
63) Chlorobenzene-d5	11.426	117	350362	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.359	152	111224	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	80259	30.692	ug/l	-0.01
Spiked Amount 50.000	Range 50 - 163		Recovery =	61.380%		
35) Dibromofluoromethane	7.634	113	72219	34.859	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	69.720%		
50) Toluene-d8	10.103	98	381179	48.592	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	97.180%		
62) 4-Bromofluorobenzene	12.420	95	131390	49.240	ug/l	0.01
Spiked Amount 50.000	Range 30 - 143		Recovery =	98.480%		
Target Compounds						Qvalue
3) Chloromethane	2.050	50	10065	3.144	ug/l	95
5) Bromomethane	2.531	94	14874	5.883	ug/l	96
10) Methyl Iodide	3.860	142	25582	9.028	ug/l	93
11) Tert butyl alcohol	4.763	59	7729	25.197	ug/l #	72
13) Acrolein	3.641	56	237	1.782	ug/l #	10
18) Methyl Acetate	4.385	43	9972	7.880	ug/l #	62
20) Methylene Chloride	4.519	84	13634	4.554	ug/l	99
31) Cyclohexane	7.652	56	10363	2.116	ug/l #	72
39) Methylcyclohexane	9.085	83	27700	6.963	ug/l	96
40) Benzene	8.061	78	83204	8.710	ug/l	99
48) Methyl methacrylate	9.225	41	2463	1.928	ug/l #	47
49) 1,4-Dioxane	9.268	88	23	1.852	ug/l #	22
51) 4-Methyl-2-Pentanone	10.030	43	3396	2.365	ug/l	90
52) Toluene	10.170	92	70079	11.636	ug/l	99
53) t-1,3-Dichloropropene	10.280	75	7998	2.702	ug/l #	43
55) 1,1,2-Trichloroethane	10.536	97	7769	4.695	ug/l #	18
56) Ethyl methacrylate	10.438	69	2239	1.061	ug/l #	1
59) 2-Hexanone	10.725	43	2414	2.543	ug/l #	32
60) Dibromochloromethane	10.646	129	2664	1.174	ug/l #	10
64) Tetrachloroethene	10.646	164	4127	1.449	ug/l	96
67) Ethyl Benzene	11.524	91	49798	3.429	ug/l	98
68) m/p-Xylenes	11.633	106	100473	18.224	ug/l	98
69) o-Xylene	11.963	106	43016	8.466	ug/l	99
73) Isopropylbenzene	12.267	105	12235	1.396	ug/l	89
76) 1,2,3-Trichloropropane	12.420	75	72113	64.341	ug/l #	100
78) n-propylbenzene	12.609	91	14148	1.330	ug/l	95
79) 2-Chlorotoluene	12.670	91	6343	1.043	ug/l #	46
80) 1,3,5-Trimethylbenzene	12.743	105	28702	4.031	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.420	75	72113	146.268	ug/l #	10
84) 1,2,4-Trimethylbenzene	13.054	105	49540	7.008	ug/l	98
95) Naphthalene	15.145	128	4032	2.785	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030725\
Data File : VY021455.D
Acq On : 07 Mar 2025 12:55
Operator : SY/MD
Sample : Q1463-03
Misc : 5.45g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
T3

Quant Time: Mar 07 22:00:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.683 min Scan# 91
Delta R.T. -0.030 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55




Instrument :
MSVOA_Y
ClientSampleId :
T3


Tgt Ion:168 Resp: 247407
Ion Ratio Lower Upper
168 100
99 55.7 46.0 69.0



#3
Chloromethane
Concen: 3.144 ug/l
RT: 2.050 min Scan# 54
Delta R.T. -0.024 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55

Tgt Ion: 50 Resp: 10065
Ion Ratio Lower Upper
50 100
52 35.7 26.2 39.2









#5
Bromomethane
Concen: 5.883 ug/l
RT: 2.531 min Scan# 11
Delta R.T. -0.073 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55




Instrument :
MSVOA_Y
ClientSampleId :
T3


Tgt Ion: 94 Resp: 14874
Ion Ratio Lower Upper
94 100
96 91.1 76.3 114.5



#10
Methyl Iodide
Concen: 9.028 ug/l
RT: 3.860 min Scan# 351
Delta R.T. -0.152 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55

Tgt Ion:142 Resp: 25582
Ion Ratio Lower Upper
142 100
127 43.9 39.1 58.7
141 11.8 11.8 17.6









#11
Tert butyl alcohol
Concen: 25.197 ug/l
RT: 4.763 min Scan# 499
Delta R.T. -0.109 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55




Instrument :
MSVOA_Y
ClientSampleId :
T3


Tgt Ion: 59 Resp: 7729
Ion Ratio Lower Upper
59 100
57 0.0 8.2 12.4#



#13
Acrolein
Concen: 1.782 ug/l
RT: 3.641 min Scan# 315
Delta R.T. -0.024 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55

Tgt Ion: 56 Resp: 237
Ion Ratio Lower Upper
56 100
55 0.0 63.4 95.2#





#18

Methyl Acetate

Concen: 7.880 ug/l

RT: 4.385 min Scan# 41

Delta R.T. -0.006 min

Lab File: VY021455.D

Acq: 07 Mar 2025 12:55

Instrument :

MSVOA_Y

ClientSampleId :





T3

Tgt Ion: 43 Resp: 9972

Ion Ratio Lower Upper

43 100

74 5.3 19.4 29.0#



#20

Methylene Chloride

Concen: 4.554 ug/l

RT: 4.519 min Scan# 459

Delta R.T. -0.103 min

Lab File: VY021455.D

Acq: 07 Mar 2025 12:55

Tgt Ion: 84 Resp: 13634




Ion Ratio Lower Upper


84 100

49 126.5 102.0 153.0

51 39.4 33.0 49.6


86 63.4 50.5 75.7








#31
Cyclohexane
Concen: 2.116 ug/l
RT: 7.652 min Scan# 91
Delta R.T. -0.055 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55

Instrument :
MSVOA_Y
ClientSampleId :
T3







Tgt Ion: 56 Resp: 10363
Ion Ratio Lower Upper
56 100
69 24.8 27.0 40.4#
84 58.2 71.1 106.7#



#33
1,2-Dichloroethane-d4
Concen: 30.692 ug/l
RT: 8.055 min Scan# 1039
Delta R.T. -0.012 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55

Tgt Ion: 65 Resp: 80259
Ion Ratio Lower Upper
65 100
67 52.6 0.0 102.8









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.603 min Scan# 1132
Delta R.T. -0.018 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55




Instrument :
MSVOA_Y
ClientSampleId :
T3


Tgt Ion:114	Resp:	315188
Ion Ratio	Lower	Upper
114	100	
63	20.4	0.0 40.8
88	15.7	0.0 30.8



#35
Dibromofluoromethane
Concen: 34.859 ug/l
RT: 7.634 min Scan# 970
Delta R.T. -0.006 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55

Tgt Ion:113	Resp:	72219
Ion Ratio	Lower	Upper
113	100	
111	106.1	82.0 123.0
192	20.7	15.9 23.9





#39

Methylcyclohexane

Concen: 6.963 ug/l

RT: 9.085 min Scan# 11

Delta R.T. -0.030 min

Lab File: VY021455.D

Acq: 07 Mar 2025 12:55

Instrument :

MSVOA_Y

ClientSampleId :

T3





Tgt Ion: 83 Resp: 27700

Ion Ratio Lower Upper

83 100

55 80.6 60.7 91.1

98 47.4 38.6 57.8



#40

Benzene

Concen: 8.710 ug/l

RT: 8.061 min Scan# 1040

Delta R.T. -0.024 min

Lab File: VY021455.D




Acq: 07 Mar 2025 12:55


Tgt Ion: 78 Resp: 83204

Ion Ratio Lower Upper

78 100

77 24.2 19.0 28.4





#48

Methyl methacrylate

Concen: 1.928 ug/l

RT: 9.225 min Scan# 1231

Delta R.T. 0.000 min

Lab File: VY021455.D

Acq: 07 Mar 2025 12:55

Instrument :

MSVOA_Y

ClientSampleId :

T3





Tgt Ion: 41 Resp: 2463

Ion Ratio Lower Upper

41 100

69 32.1 67.8 101.8#

39 23.8 46.4 69.6#



#49

1,4-Dioxane

Concen: 1.852 ug/l

RT: 9.268 min Scan# 1238

Delta R.T. 0.037 min

Lab File: VY021455.D

Acq: 07 Mar 2025 12:55




Tgt Ion: 88 Resp: 23


Ion Ratio Lower Upper

88 100

43 0.0 25.0 37.4#

58 0.0 58.4 87.6#





#50

Toluene-d8

Concen: 48.592 ug/l

RT: 10.103 min Scan# 1375

Delta R.T. -0.006 min

Lab File: VY021455.D

Acq: 07 Mar 2025 12:55

Instrument :

MSVOA_Y

ClientSampleId :





T3

Tgt Ion: 98 Resp: 381179

Ion Ratio Lower Upper

98 100

100 64.0 52.1 78.1



#51

4-Methyl-2-Pentanone

Concen: 2.365 ug/l

RT: 10.030 min Scan# 1363

Delta R.T. 0.031 min

Lab File: VY021455.D




Acq: 07 Mar 2025 12:55


Tgt Ion: 43 Resp: 3396

Ion Ratio Lower Upper

43 100

58 33.0 31.4 47.2









#52
Toluene
Concen: 11.636 ug/l
RT: 10.170 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55




Instrument :
MSVOA_Y
ClientSampleId :
T3


Tgt Ion: 92 Resp: 70079
Ion Ratio Lower Upper
92 100
91 170.8 138.2 207.2



#53
t-1,3-Dichloropropene
Concen: 2.702 ug/l
RT: 10.280 min Scan# 1404
Delta R.T. -0.115 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55

Tgt Ion: 75 Resp: 7998
Ion Ratio Lower Upper
75 100
77 0.0 25.0 37.6#





#55

1,1,2-Trichloroethane

Concen: 4.695 ug/l

RT: 10.536 min Scan# 1446

Delta R.T. -0.036 min

Lab File: VY021455.D

Acq: 07 Mar 2025 12:55

Instrument :

MSVOA_Y

ClientSampleId :

T3

Tgt Ion: 97 Resp: 7769





Ion Ratio Lower Upper

97 100

83 5.4 70.0 105.0#

85 3.0 44.3 66.5#

99 0.0 51.4 77.0#



#56

Ethyl methacrylate

Concen: 1.061 ug/l

RT: 10.438 min Scan# 1430

Delta R.T. -0.006 min

Lab File: VY021455.D

Acq: 07 Mar 2025 12:55




Tgt Ion: 69 Resp: 2239


Ion Ratio Lower Upper

69 100

41 184.9 53.3 79.9#

39 83.7 31.0 46.6#





#59

2-Hexanone

Concen: 2.543 ug/l

RT: 10.725 min Scan# 14

Delta R.T. -0.036 min

Lab File: VY021455.D

Acq: 07 Mar 2025 12:55

Instrument :

MSVOA_Y

ClientSampleId :





T3

Tgt Ion: 43 Resp: 2414

Ion Ratio Lower Upper

43 100

58 5.8 27.3 81.8#



#60

Dibromochloromethane

Concen: 1.174 ug/l

RT: 10.646 min Scan# 1464

Delta R.T. -0.268 min

Lab File: VY021455.D




Acq: 07 Mar 2025 12:55


Tgt Ion: 129 Resp: 2664

Ion Ratio Lower Upper

129 100

127 0.0 39.1 117.3#









#62
4-Bromofluorobenzene
Concen: 49.240 ug/l
RT: 12.420 min Scan# 1755
Delta R.T. 0.013 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55




Instrument :
MSVOA_Y
ClientSampleId :
T3


Tgt Ion: 95 Resp: 131390
Ion Ratio Lower Upper
95 100
174 84.0 0.0 165.0
176 82.0 0.0 160.0



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.426 min Scan# 1592
Delta R.T. 0.006 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55

Tgt Ion: 117 Resp: 350362
Ion Ratio Lower Upper
117 100
82 55.3 44.6 67.0
119 33.6 25.4 38.0









#64
Tetrachloroethene
Concen: 1.449 ug/l
RT: 10.646 min Scan# 1465
Delta R.T. -0.006 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55




Instrument :
MSVOA_Y
ClientSampleId :
T3


Tgt Ion:164	Resp:	4127
Ion	Ratio	Lower Upper
164	100	
166	127.0	100.4 150.6
129	96.5	73.8 110.8
131	82.9	71.5 107.3



#67
Ethyl Benzene
Concen: 3.429 ug/l
RT: 11.524 min Scan# 1608
Delta R.T. 0.000 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55

Tgt Ion: 91	Resp:	49798
Ion	Ratio	Lower Upper
91	100	
106	29.7	24.6 36.8









#68
m/p-Xylenes
Concen: 18.224 ug/l
RT: 11.633 min Scan# 1
Delta R.T. 0.000 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55




Instrument :
MSVOA_Y
ClientSampleId :
T3


Tgt Ion:106 Resp: 100473
Ion Ratio Lower Upper
106 100
91 204.4 161.7 242.5



#69
o-Xylene
Concen: 8.466 ug/l
RT: 11.963 min Scan# 1680
Delta R.T. 0.006 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55

Tgt Ion:106 Resp: 43016
Ion Ratio Lower Upper
106 100
91 216.2 107.6 322.8









#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.359 min Scan# 1908
Delta R.T. 0.006 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55




Instrument :
MSVOA_Y
ClientSampleId :
T3


Tgt Ion:152 Resp: 111224
Ion Ratio Lower Upper
152 100
115 58.9 29.0 87.0
150 159.0 0.0 347.2



#73
Isopropylbenzene
Concen: 1.396 ug/l
RT: 12.267 min Scan# 1730
Delta R.T. 0.013 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55

Tgt Ion:105 Resp: 12235
Ion Ratio Lower Upper
105 100
120 20.5 13.1 39.3









#76
1,2,3-Trichloropropane
Concen: 64.341 ug/l
RT: 12.420 min Scan# 1755
Delta R.T. -0.140 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55




Instrument :
MSVOA_Y
ClientSampleId :
T3


Tgt Ion: 75 Resp: 72113
Ion Ratio Lower Upper
75 100
77 1.5 0.0 0.0#



#78
n-propylbenzene
Concen: 1.330 ug/l
RT: 12.609 min Scan# 1786
Delta R.T. 0.013 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55

Tgt Ion: 91 Resp: 14148
Ion Ratio Lower Upper
91 100
120 24.7 11.3 33.8





#79

2-Chlorotoluene

Concen: 1.043 ug/l

RT: 12.670 min Scan# 1796

Delta R.T. -0.012 min

Lab File: VY021455.D

Acq: 07 Mar 2025 12:55

Instrument :

MSVOA_Y

ClientSampleId :





T3

Tgt Ion: 91 Resp: 6343

Ion Ratio Lower Upper

91 100

126 3.6 17.2 51.6#



#80

1,3,5-Trimethylbenzene

Concen: 4.031 ug/l

RT: 12.743 min Scan# 1808

Delta R.T. 0.000 min

Lab File: VY021455.D




Acq: 07 Mar 2025 12:55


Tgt Ion: 105 Resp: 28702

Ion Ratio Lower Upper

105 100

120 50.5 24.6 73.6









#81
trans-1,4-Dichloro-2-butene
Concen: 146.268 ug/l
RT: 12.420 min Scan# 1755
Delta R.T. 0.116 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55




Instrument :
MSVOA_Y
ClientSampleId :
T3


Tgt Ion: 75 Resp: 72113
Ion Ratio Lower Upper
75 100
53 0.3 81.0 121.4#
89 0.3 36.9 55.3#



#84
1,2,4-Trimethylbenzene
Concen: 7.008 ug/l
RT: 13.054 min Scan# 1859
Delta R.T. 0.006 min
Lab File: VY021455.D
Acq: 07 Mar 2025 12:55

Tgt Ion: 105 Resp: 49540
Ion Ratio Lower Upper
105 100
120 44.2 22.7 68.1





#95

Naphthalene

Concen: 2.785 ug/l

RT: 15.145 min Scan# 21

Delta R.T. -0.006 min

Lab File: VY021455.D


Acq: 07 Mar 2025 12:55

Instrument :

MSVOA_Y

ClientSampleId :

T3



Tgt Ion:128 Resp: 4032

Ion Ratio Lower Upper

128 100

127 15.5 10.6 15.8

129 12.9 8.6 13.0

