

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030725\
 Data File : VY021453.D
 Acq On : 07 Mar 2025 12:03
 Operator : SY/MD
 Sample : VY0307SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0307SBS01

Quant Time: Mar 07 22:00:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 03/11/2025
 Supervised By :Semsettin Yesilyurt 03/11/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	219868	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	343563	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	301562	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	152803	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	123341	53.076	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 106.160%			
35) Dibromofluoromethane	7.634	113	117478	52.021	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 104.040%			
50) Toluene-d8	10.109	98	444032	51.930	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 103.860%			
62) 4-Bromofluorobenzene	12.408	95	148084	50.913	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 101.820%			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.867	85	39418	19.537	ug/l	94
3) Chloromethane	2.068	50	57411	20.182	ug/l	98
4) Vinyl Chloride	2.202	62	64302	20.685	ug/l	96
5) Bromomethane	2.592	94	46339	20.625	ug/l	96
6) Chloroethane	2.732	64	44175	21.504	ug/l	97
7) Trichlorofluoromethane	3.056	101	94934	21.880	ug/l	95
8) Diethyl Ether	3.452	74	23849	19.834	ug/l	99
9) 1,1,2-Trichlorotrifluo...	3.818	101	49349	19.921	ug/l	99
10) Methyl Iodide	4.007	142	48739	19.355	ug/l	100
11) Tert butyl alcohol	4.872	59	18258	108.044	ug/l	99
12) 1,1-Dichloroethene	3.787	96	43870	19.424	ug/l	97
13) Acrolein	3.647	56	16580	140.270	ug/l	95
14) Allyl chloride	4.385	41	70258	19.249	ug/l	99
15) Acrylonitrile	5.061	53	50908	99.456	ug/l	98
16) Acetone	3.866	43	58595	115.604	ug/l	96
17) Carbon Disulfide	4.104	76	133704	18.890	ug/l	99
18) Methyl Acetate	4.385	43	23025	20.473	ug/l	100
19) Methyl tert-butyl Ether	5.116	73	110998	19.390	ug/l	99
20) Methylene Chloride	4.616	84	54760	20.582	ug/l	95
21) trans-1,2-Dichloroethene	5.116	96	49678	19.641	ug/l	95
22) Diisopropyl ether	6.018	45	157929	20.078	ug/l	99
23) Vinyl Acetate	5.957	43	450639	100.386	ug/l	99
24) 1,1-Dichloroethane	5.915	63	91407	19.690	ug/l	99
25) 2-Butanone	6.896	43	75161	107.199	ug/l	99
26) 2,2-Dichloropropane	6.884	77	82994	19.429	ug/l	98
27) cis-1,2-Dichloroethene	6.896	96	54488	18.963	ug/l	98
28) Bromochloromethane	7.244	49	37442	19.238	ug/l	99
29) Tetrahydrofuran	7.262	42	44158	102.018	ug/l	97
30) Chloroform	7.421	83	95944	19.903	ug/l	99
31) Cyclohexane	7.701	56	82128	18.870	ug/l	97
32) 1,1,1-Trichloroethane	7.616	97	86978	19.705	ug/l	99
36) 1,1-Dichloropropene	7.835	75	68637	19.417	ug/l	98
37) Ethyl Acetate	6.982	43	31592	20.220	ug/l	97
38) Carbon Tetrachloride	7.817	117	79070	19.449	ug/l	96
39) Methylcyclohexane	9.109	83	81593	18.818	ug/l	99
40) Benzene	8.079	78	202793	19.475	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030725\
 Data File : VY021453.D
 Acq On : 07 Mar 2025 12:03
 Operator : SY/MD
 Sample : VY0307SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0307SBS01

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 03/11/2025
 Supervised By :Semsettin Yesilyurt 03/11/2025

Quant Time: Mar 07 22:00:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.226	41	17773m	20.892	ug/l	
42) 1,2-Dichloroethane	8.158	62	57755	19.826	ug/l	100
43) Isopropyl Acetate	8.195	43	59917	19.755	ug/l #	87
44) Trichloroethene	8.865	130	50339	19.130	ug/l	99
45) 1,2-Dichloropropane	9.140	63	49619	20.021	ug/l	97
46) Dibromomethane	9.231	93	27131	19.675	ug/l	99
47) Bromodichloromethane	9.426	83	71823	19.613	ug/l	99
48) Methyl methacrylate	9.219	41	27630	19.839	ug/l	98
49) 1,4-Dioxane	9.237	88	5803	428.740	ug/l	95
51) 4-Methyl-2-Pentanone	9.999	43	157126	100.404	ug/l	98
52) Toluene	10.170	92	128919	19.637	ug/l	98
53) t-1,3-Dichloropropene	10.396	75	62838	19.476	ug/l	97
54) cis-1,3-Dichloropropene	9.859	75	73636	19.223	ug/l	96
55) 1,1,2-Trichloroethane	10.573	97	35572	19.720	ug/l	96
56) Ethyl methacrylate	10.438	69	44620	19.396	ug/l	100
57) 1,3-Dichloropropane	10.719	76	60936	19.491	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.713	63	107568	101.290	ug/l	99
59) 2-Hexanone	10.762	43	109553	105.872	ug/l	98
60) Dibromochloromethane	10.914	129	47870	19.358	ug/l	99
61) 1,2-Dibromoethane	11.018	107	33437	19.641	ug/l	97
64) Tetrachloroethene	10.646	164	48058	19.598	ug/l	98
65) Chlorobenzene	11.444	112	135015	18.959	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.517	131	48814	19.388	ug/l	100
67) Ethyl Benzene	11.517	91	238496	19.081	ug/l	99
68) m/p-Xylenes	11.627	106	184669	38.916	ug/l	99
69) o-Xylene	11.956	106	83621	19.120	ug/l	99
70) Styrene	11.969	104	140205	19.324	ug/l	99
71) Bromoform	12.133	173	28337	20.134	ug/l #	98
73) Isopropylbenzene	12.255	105	227722	18.919	ug/l	100
74) N-amyl acetate	12.072	43	49116	18.789	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.505	83	41732	19.647	ug/l	100
76) 1,2,3-Trichloropropane	12.554	75	30660m	19.912	ug/l	
77) Bromobenzene	12.536	156	51851	18.413	ug/l	98
78) n-propylbenzene	12.597	91	280785	19.208	ug/l	100
79) 2-Chlorotoluene	12.682	91	159129	19.038	ug/l	99
80) 1,3,5-Trimethylbenzene	12.737	105	187183	19.136	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.304	75	13174	19.450	ug/l	95
82) 4-Chlorotoluene	12.779	91	166239	19.206	ug/l	100
83) tert-Butylbenzene	12.999	119	163102	18.691	ug/l	98
84) 1,2,4-Trimethylbenzene	13.048	105	185591	19.109	ug/l	100
85) sec-Butylbenzene	13.176	105	249181	19.192	ug/l	100
86) p-Isopropyltoluene	13.292	119	202582	18.939	ug/l	100
87) 1,3-Dichlorobenzene	13.292	146	105372	18.798	ug/l	99
88) 1,4-Dichlorobenzene	13.371	146	104875	19.131	ug/l	98
89) n-Butylbenzene	13.621	91	186240	18.856	ug/l	99
90) Hexachloroethane	13.883	117	43553	19.087	ug/l	98
91) 1,2-Dichlorobenzene	13.663	146	91257	18.995	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.273	75	6311	19.835	ug/l	90
93) 1,2,4-Trichlorobenzene	14.925	180	47173	17.977	ug/l	97
94) Hexachlorobutadiene	15.029	225	31456	19.123	ug/l	99
95) Naphthalene	15.151	128	74806	18.887	ug/l	100
96) 1,2,3-Trichlorobenzene	15.334	180	39961	18.107	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030725\
Data File : VY021453.D
Acq On : 07 Mar 2025 12:03
Operator : SY/MD
Sample : VY0307SBS01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025

Quant Time: Mar 07 22:00:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030725\
 Data File : VY021453.D
 Acq On : 07 Mar 2025 12:03
 Operator : SY/MD
 Sample : VY0307SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0307SBS01

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 03/11/2025
 Supervised By :Semsettin Yesilyurt 03/11/2025


Quant Time: Mar 07 22:00:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01



Tgt Ion:168 Resp: 21986
Ion Ratio Lower Upper
168 100
99 56.3 46.0 69.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#2
Dichlorodifluoromethane
Concen: 19.537 ug/l
RT: 1.867 min Scan# 24
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03

Tgt Ion: 85 Resp: 39418
Ion Ratio Lower Upper
85 100
87 34.3 15.7 47.0









#3
Chloromethane
Concen: 20.182 ug/l
RT: 2.068 min Scan# 51
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03

Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01

Tgt Ion: 50 Resp: 5741
Ion Ratio Lower Upper
50 100
52 31.7 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#4
Vinyl Chloride
Concen: 20.685 ug/l
RT: 2.202 min Scan# 79
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03


Tgt Ion: 62 Resp: 64302
Ion Ratio Lower Upper
62 100
64 30.5 26.2 39.2





#5
Bromomethane
Concen: 20.625 ug/l
RT: 2.592 min Scan# 145
Delta R.T. -0.012 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01




Tgt Ion: 94 Resp: 46339
Ion Ratio Lower Upper
94 100
96 91.7 76.3 114.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025




#6
Chloroethane
Concen: 21.504 ug/l
RT: 2.732 min Scan# 166
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Tgt Ion: 64 Resp: 44175
Ion Ratio Lower Upper
64 100
66 31.6 26.7 40.1





#7
Trichlorofluoromethane
Concen: 21.880 ug/l
RT: 3.056 min Scan# 21
Delta R.T. -0.012 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01




Tgt Ion:101 Resp: 94934
Ion Ratio Lower Upper
101 100
103 61.0 51.7 77.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025




#8
Diethyl Ether
Concen: 19.834 ug/l
RT: 3.452 min Scan# 284
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03



Tgt Ion: 74 Resp: 23849
Ion Ratio Lower Upper
74 100
45 100.2 49.7 149.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.921 ug/l
RT: 3.818 min Scan# 346
Delta R.T. -0.012 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03

Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01

Tgt Ion:101 Resp: 49349
Ion Ratio Lower Upper
101 100
85 47.2 37.1 55.7
151 76.4 62.1 93.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#10
Methyl Iodide
Concen: 19.355 ug/l
RT: 4.007 min Scan# 375
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03


Tgt Ion:142 Resp: 48739
Ion Ratio Lower Upper
142 100
127 48.6 39.1 58.7
141 14.6 11.8 17.6





#11
Tert butyl alcohol
Concen: 108.044 ug/l
RT: 4.872 min Scan# 517
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01




Tgt Ion: 59 Resp: 1825
Ion Ratio Lower Upper
59 100
57 9.8 8.2 12.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025




#12
1,1-Dichloroethene
Concen: 19.424 ug/l
RT: 3.787 min Scan# 339
Delta R.T. -0.012 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Tgt Ion: 96 Resp: 43870
Ion Ratio Lower Upper
96 100
61 174.1 135.2 202.8
98 62.9 51.4 77.2





#13
Acrolein
Concen: 140.270 ug/l
RT: 3.647 min Scan# 319
Delta R.T. -0.018 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01



Tgt Ion: 56 Resp: 16580
Ion Ratio Lower Upper
56 100
55 74.8 63.4 95.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#14
Allyl chloride
Concen: 19.249 ug/l
RT: 4.385 min Scan# 437
Delta R.T. -0.012 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03


Tgt Ion: 41 Resp: 70258
Ion Ratio Lower Upper
41 100
39 74.4 59.3 88.9
76 37.8 29.8 44.8





#15
Acrylonitrile
Concen: 99.456 ug/l
RT: 5.061 min Scan# 549
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01



Tgt Ion: 53 Resp: 50908
Ion Ratio Lower Upper
53 100
52 82.2 66.3 99.5
51 40.1 30.3 45.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#16
Acetone
Concen: 115.604 ug/l
RT: 3.866 min Scan# 352
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03


Tgt Ion: 43 Resp: 58595
Ion Ratio Lower Upper
43 100
58 29.8 22.1 33.1





#17
Carbon Disulfide
Concen: 18.890 ug/l
RT: 4.104 min Scan# 391
Delta R.T. -0.012 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01



Tgt Ion: 76 Resp: 133704
Ion Ratio Lower Upper
76 100
78 9.5 7.4 11.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#18
Methyl Acetate
Concen: 20.473 ug/l
RT: 4.385 min Scan# 437
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03


Tgt Ion: 43 Resp: 23025
Ion Ratio Lower Upper
43 100
74 24.0 19.4 29.0





#19
Methyl tert-butyl Ether
Concen: 19.390 ug/l
RT: 5.116 min Scan# 51
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01



Tgt Ion: 73 Resp: 110998
Ion Ratio Lower Upper
73 100
57 21.7 17.9 26.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#20
Methylene Chloride
Concen: 20.582 ug/l
RT: 4.616 min Scan# 475
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03


Tgt Ion: 84 Resp: 54760
Ion Ratio Lower Upper
84 100
49 121.2 102.0 153.0
51 38.3 33.0 49.6
86 59.6 50.5 75.7





#21
trans-1,2-Dichloroethene
Concen: 19.641 ug/l
RT: 5.116 min Scan# 51
Delta R.T. -0.012 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01



Tgt Ion: 96 Resp: 49678
Ion Ratio Lower Upper
96 100
61 145.0 111.5 167.3
98 61.9 53.8 80.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#22
Diisopropyl ether
Concen: 20.078 ug/l
RT: 6.018 min Scan# 705
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03

Tgt Ion: 45 Resp: 157929
Ion Ratio Lower Upper
45 100
43 61.6 48.7 73.1
87 26.7 21.5 32.3
59 11.4 9.3 13.9





#23

Vinyl Acetate

Concen: 100.386 ug/l

RT: 5.957 min Scan# 64

Delta R.T. -0.012 min

Lab File: VY021453.D

Acq: 07 Mar 2025 12:03

Instrument :

MSVOA_Y

ClientSampleId :

VY0307SBS01

Tgt Ion: 43 Resp: 450639

Ion Ratio Lower Upper

43 100





86 10.1 8.5 12.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/11/2025

Supervised By :Semsettin Yesilyurt 03/11/2025



#24

1,1-Dichloroethane

Concen: 19.690 ug/l

RT: 5.915 min Scan# 688

Delta R.T. -0.012 min

Lab File: VY021453.D

Acq: 07 Mar 2025 12:03




Tgt Ion: 63 Resp: 91407


Ion Ratio Lower Upper

63 100

98 6.7 3.3 9.9

100 5.0 2.1 6.5





#25

2-Butanone

Concen: 107.199 ug/l

RT: 6.896 min Scan# 849

Delta R.T. 0.000 min

Lab File: VY021453.D

Acq: 07 Mar 2025 12:03

Instrument :

MSVOA_Y

ClientSampleId :

VY0307SBS01

Tgt Ion: 43 Resp: 7516

Ion Ratio Lower Upper

43 100





72 26.0 21.4 32.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/11/2025

Supervised By :Semsettin Yesilyurt 03/11/2025



#26

2,2-Dichloropropane

Concen: 19.429 ug/l

RT: 6.884 min Scan# 847

Delta R.T. -0.006 min

Lab File: VY021453.D




Acq: 07 Mar 2025 12:03


Tgt Ion: 77 Resp: 82994

Ion Ratio Lower Upper

77 100


97 21.7 11.3 33.8





#27
 cis-1,2-Dichloroethene
 Concen: 18.963 ug/l
 RT: 6.896 min Scan# 849
 Delta R.T. 0.000 min
 Lab File: VY021453.D
 Acq: 07 Mar 2025 12:03




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0307SBS01



Tgt Ion: 96 Resp: 5448
 Ion Ratio Lower Upper
 96 100
 61 152.5 0.0 300.8
 98 66.5 0.0 130.0




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
 Supervised By :Semsettin Yesilyurt 03/11/2025



#28
 Bromochloromethane
 Concen: 19.238 ug/l
 RT: 7.244 min Scan# 906
 Delta R.T. -0.006 min
 Lab File: VY021453.D
 Acq: 07 Mar 2025 12:03

Tgt Ion: 49 Resp: 37442
 Ion Ratio Lower Upper
 49 100
 129 2.3 0.0 4.2
 130 75.4 59.6 89.4





#29

Tetrahydrofuran

Concen: 102.018 ug/l

RT: 7.262 min Scan# 909

Delta R.T. -0.006 min

Lab File: VY021453.D

Acq: 07 Mar 2025 12:03

Instrument :

MSVOA_Y

ClientSampleId :





VY0307SBS01

Tgt Ion:	42	Resp:	44158
Ion Ratio	Lower	Upper	
42	100		
72	41.5	34.8	52.2
71	39.3	32.6	48.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#30

Chloroform

Concen: 19.903 ug/l




RT: 7.421 min Scan# 935


Delta R.T. -0.006 min

Lab File: VY021453.D

Acq: 07 Mar 2025 12:03


Tgt Ion:	83	Resp:	95944
Ion Ratio	Lower	Upper	
83	100		
85	65.6	53.1	79.7





#31
Cyclohexane
Concen: 18.870 ug/l
RT: 7.701 min Scan# 982
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01



Tgt Ion: 56 Resp: 82128
Ion Ratio Lower Upper
56 100
69 35.6 27.0 40.4
84 86.6 71.1 106.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#32
1,1,1-Trichloroethane
Concen: 19.705 ug/l
RT: 7.616 min Scan# 967
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03

Tgt Ion: 97 Resp: 86978
Ion Ratio Lower Upper
97 100
99 64.5 51.7 77.5
61 45.6 35.7 53.5









#33
1,2-Dichloroethane-d4
Concen: 53.076 ug/l
RT: 8.061 min Scan# 1104
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03

Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01

Tgt Ion: 65 Resp: 12334
Ion Ratio Lower Upper
65 100
67 51.8 0.0 102.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.616 min Scan# 1131
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03


Tgt Ion:114 Resp: 343563
Ion Ratio Lower Upper
114 100
63 21.8 0.0 40.8
88 15.4 0.0 30.8





#35
Dibromofluoromethane
Concen: 52.021 ug/l
RT: 7.634 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01




Tgt Ion:113 Resp: 117478
Ion Ratio Lower Upper
113 100
111 101.6 82.0 123.0
192 19.9 15.9 23.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025




#36
1,1-Dichloropropene
Concen: 19.417 ug/l
RT: 7.835 min Scan# 1003
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03



Tgt Ion: 75 Resp: 68637
Ion Ratio Lower Upper
75 100
110 33.6 17.3 51.7
77 30.7 25.4 38.2





#37

Ethyl Acetate

Concen: 20.220 ug/l

RT: 6.982 min Scan# 81

Delta R.T. -0.006 min

Lab File: VY021453.D

Acq: 07 Mar 2025 12:03

Instrument :

MSVOA_Y

ClientSampleId :

VY0307SBS01





Tgt Ion: 43 Resp: 31592

Ion	Ratio	Lower	Upper
43	100		
61	12.5	11.4	17.0
70	9.5	8.2	12.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#38

Carbon Tetrachloride

Concen: 19.449 ug/l

RT: 7.817 min Scan# 1000




Delta R.T. -0.006 min


Lab File: VY021453.D

Acq: 07 Mar 2025 12:03

Tgt Ion:117 Resp: 79070


Ion	Ratio	Lower	Upper
117	100		
119	99.0	75.8	113.8
121	31.3	23.9	35.9





#39
Methylcyclohexane
Concen: 18.818 ug/l
RT: 9.109 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01



Tgt Ion: 83 Resp: 8159
Ion Ratio Lower Upper
83 100
55 75.4 60.7 91.1
98 48.7 38.6 57.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#40
Benzene
Concen: 19.475 ug/l
RT: 8.079 min Scan# 1043
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03


Tgt Ion: 78 Resp: 202793
Ion Ratio Lower Upper
78 100
77 23.1 19.0 28.4





#41
Methacrylonitrile
Concen: 20.892 ug/l m
RT: 7.226 min Scan# 902
Delta R.T. 0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01



Tgt Ion: 41 Resp: 1777
Ion Ratio Lower Upper
41 100
39 63.6 35.0 52.6
67 0.0 61.2 91.8
52 0.0 26.8 40.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#42
1,2-Dichloroethane
Concen: 19.826 ug/l
RT: 8.158 min Scan# 1056
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03

Tgt Ion: 62 Resp: 57755
Ion Ratio Lower Upper
62 100
98 9.2 0.0 18.2





#43

Isopropyl Acetate

Concen: 19.755 ug/l

RT: 8.195 min Scan# 1062

Delta R.T. -0.006 min

Lab File: VY021453.D

Acq: 07 Mar 2025 12:03

Instrument :

MSVOA_Y

ClientSampleId :

VY0307SBS01





Tgt Ion:	43	Resp:	5991
Ion Ratio	Lower	Upper	
43	100		
61	19.8	23.4	35.0
87	12.7	10.5	15.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/11/2025

Supervised By :Semsettin Yesilyurt 03/11/2025



#44

Trichloroethene

Concen: 19.130 ug/l




RT: 8.865 min Scan# 1172


Delta R.T. -0.006 min

Lab File: VY021453.D

Acq: 07 Mar 2025 12:03

Tgt Ion:	130	Resp:	50339
Ion Ratio	Lower	Upper	
130	100		
95	96.6	0.0	191.6





#45

1,2-Dichloropropane

Concen: 20.021 ug/l

RT: 9.140 min Scan# 1217

Delta R.T. -0.006 min

Lab File: VY021453.D

Acq: 07 Mar 2025 12:03

Instrument :

MSVOA_Y

ClientSampleId :

VY0307SBS01

Tgt Ion: 63 Resp: 49619

Ion Ratio Lower Upper

63 100





65 30.7 26.0 39.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/11/2025

Supervised By :Semsettin Yesilyurt 03/11/2025



#46

Dibromomethane

Concen: 19.675 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. -0.006 min

Lab File: VY021453.D

Acq: 07 Mar 2025 12:03




Tgt Ion: 93 Resp: 27131


Ion Ratio Lower Upper

93 100

95 80.8 65.8 98.8

174 97.5 77.9 116.9









#47
Bromodichloromethane
Concen: 19.613 ug/l
RT: 9.426 min Scan# 1264
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03

Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01

Tgt Ion: 83 Resp: 7182
Ion Ratio Lower Upper
83 100
85 63.4 51.1 76.7
127 9.5 6.7 10.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#48
Methyl methacrylate
Concen: 19.839 ug/l
RT: 9.219 min Scan# 1230
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03


Tgt Ion: 41 Resp: 27630
Ion Ratio Lower Upper
41 100
69 84.0 67.8 101.8
39 55.6 46.4 69.6





#49
1,4-Dioxane
Concen: 428.740 ug/l
RT: 9.237 min Scan# 1233
Delta R.T. 0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01



Tgt Ion: 88 Resp: 580
Ion Ratio Lower Upper
88 100
43 34.8 25.0 37.4
58 77.2 58.4 87.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#50
Toluene-d8
Concen: 51.930 ug/l
RT: 10.109 min Scan# 1376
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03

Tgt Ion: 98 Resp: 444032
Ion Ratio Lower Upper
98 100
100 64.8 52.1 78.1









#51
4-Methyl-2-Pentanone
Concen: 100.404 ug/l
RT: 9.999 min Scan# 1358
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03

Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01

Tgt Ion: 43 Resp: 157120
Ion Ratio Lower Upper
43 100
58 38.2 31.4 47.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#52
Toluene
Concen: 19.637 ug/l
RT: 10.170 min Scan# 1386
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03

Tgt Ion: 92 Resp: 128919
Ion Ratio Lower Upper
92 100
91 170.4 138.2 207.2









#53
t-1,3-Dichloropropene
Concen: 19.476 ug/l
RT: 10.396 min Scan# 1423
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03

Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01

Tgt Ion: 75 Resp: 62838
Ion Ratio Lower Upper
75 100
77 32.7 25.0 37.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#54
cis-1,3-Dichloropropene
Concen: 19.223 ug/l
RT: 9.859 min Scan# 1335
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03


Tgt Ion: 75 Resp: 73636
Ion Ratio Lower Upper
75 100
77 28.9 24.6 37.0
39 45.8 39.4 59.2





#55
1,1,2-Trichloroethane
Concen: 19.720 ug/l
RT: 10.573 min Scan# 1452
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01




Tgt Ion	Ratio	Lower	Upper
97	100		
83	88.8	70.0	105.0
85	53.8	44.3	66.5
99	57.9	51.4	77.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025




#56
Ethyl methacrylate
Concen: 19.396 ug/l
RT: 10.438 min Scan# 1430
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03



Tgt Ion	Ratio	Lower	Upper
69	100		
41	66.9	53.3	79.9
39	38.7	31.0	46.6





#57

1,3-Dichloropropane

Concen: 19.491 ug/l

RT: 10.719 min Scan# 1476

Delta R.T. 0.000 min

Lab File: VY021453.D

Acq: 07 Mar 2025 12:03

Instrument :

MSVOA_Y

ClientSampleId :

VY0307SBS01

Tgt Ion: 76 Resp: 60930

Ion Ratio Lower Upper

76 100





78 32.3 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/11/2025

Supervised By :Semsettin Yesilyurt 03/11/2025



#58

2-Chloroethyl Vinyl ether

Concen: 101.290 ug/l

RT: 9.713 min Scan# 1311

Delta R.T. 0.000 min

Lab File: VY021453.D




Acq: 07 Mar 2025 12:03


Tgt Ion: 63 Resp: 107568

Ion Ratio Lower Upper

63 100


106 27.3 22.1 33.1





#59
2-Hexanone
Concen: 105.872 ug/l
RT: 10.762 min Scan# 1483
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01



Tgt Ion: 43 Resp: 10955
Ion Ratio Lower Upper
43 100
58 52.7 27.3 81.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#60
Dibromochloromethane
Concen: 19.358 ug/l
RT: 10.914 min Scan# 1508
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03


Tgt Ion:129 Resp: 47870
Ion Ratio Lower Upper
129 100
127 79.4 39.1 117.3





#61
1,2-Dibromoethane
Concen: 19.641 ug/l
RT: 11.018 min Scan# 1525
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01




Tgt Ion: 107 Resp: 3343
Ion Ratio Lower Upper
107 100
109 92.3 75.8 113.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025




#62
4-Bromofluorobenzene
Concen: 50.913 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Tgt Ion: 95 Resp: 148084
Ion Ratio Lower Upper
95 100
174 82.8 0.0 165.0
176 79.4 0.0 160.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.420 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01



Tgt Ion:117 Resp: 301562
Ion Ratio Lower Upper
117 100
82 56.5 44.6 67.0
119 33.0 25.4 38.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#64
Tetrachloroethene
Concen: 19.598 ug/l
RT: 10.646 min Scan# 1464
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03


Tgt Ion:164 Resp: 48058
Ion Ratio Lower Upper
164 100
166 122.2 100.4 150.6
129 89.9 73.8 110.8
131 90.7 71.5 107.3





#65
Chlorobenzene
Concen: 18.959 ug/l
RT: 11.444 min Scan# 1595
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01



Tgt Ion:112 Resp: 135015
Ion Ratio Lower Upper
112 100
114 32.5 26.1 39.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 19.388 ug/l
RT: 11.517 min Scan# 1607
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03

Tgt Ion:131 Resp: 48814
Ion Ratio Lower Upper
131 100
133 94.7 47.6 142.9
119 64.8 32.3 96.9





#67

Ethyl Benzene

Concen: 19.081 ug/l

RT: 11.517 min Scan# 1607

Delta R.T. -0.006 min

Lab File: VY021453.D

Acq: 07 Mar 2025 12:03

Instrument :

MSVOA_Y

ClientSampleId :

VY0307SBS01

Tgt Ion: 91 Resp: 238490

Ion Ratio Lower Upper

91 100





106 31.2 24.6 36.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/11/2025

Supervised By :Semsettin Yesilyurt 03/11/2025



#68

m/p-Xylenes

Concen: 38.916 ug/l

RT: 11.627 min Scan# 1625

Delta R.T. -0.006 min

Lab File: VY021453.D




Acq: 07 Mar 2025 12:03


Tgt Ion:106 Resp: 184669

Ion Ratio Lower Upper

106 100

91 201.0 161.7 242.5









#69
o-Xylene
Concen: 19.120 ug/l
RT: 11.956 min Scan# 1679
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03

Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01

Tgt Ion:106 Resp: 8362
Ion Ratio Lower Upper
106 100
91 216.1 107.6 322.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#70
Styrene
Concen: 19.324 ug/l
RT: 11.969 min Scan# 1681
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03


Tgt Ion:104 Resp: 140205
Ion Ratio Lower Upper
104 100
78 50.1 40.7 61.1
103 54.9 44.1 66.1





#71
Bromoform
Concen: 20.134 ug/l
RT: 12.133 min Scan# 1708
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01




Tgt Ion:173 Resp: 2833
Ion Ratio Lower Upper
173 100
175 46.6 23.9 71.7
254 0.0 0.1 0.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.346 min Scan# 1907
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Tgt Ion:152 Resp: 152803
Ion Ratio Lower Upper
152 100
115 59.3 29.0 87.0
150 163.4 0.0 347.2





#73
Isopropylbenzene
Concen: 18.919 ug/l
RT: 12.255 min Scan# 1728
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01



Tgt Ion: 105 Resp: 22772
Ion Ratio Lower Upper
105 100
120 26.4 13.1 39.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#74
N-ethyl acetate
Concen: 18.789 ug/l
RT: 12.072 min Scan# 1698
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03


Tgt Ion: 43 Resp: 49116
Ion Ratio Lower Upper
43 100
70 43.6 34.6 51.8
55 28.4 21.1 31.7
61 24.2 19.4 29.0





#75
 1,1,2,2-Tetrachloroethane
 Concen: 19.647 ug/l
 RT: 12.505 min Scan# 1769
 Delta R.T. -0.006 min
 Lab File: VY021453.D
 Acq: 07 Mar 2025 12:03




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0307SBS01



Tgt Ion: 83 Resp: 41732
 Ion Ratio Lower Upper
 83 100
 131 10.0 5.3 16.1
 85 63.6 31.9 95.5




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
 Supervised By :Semsettin Yesilyurt 03/11/2025



#76
 1,2,3-Trichloropropane
 Concen: 19.912 ug/l m
 RT: 12.554 min Scan# 1777
 Delta R.T. -0.006 min
 Lab File: VY021453.D
 Acq: 07 Mar 2025 12:03


Tgt Ion: 75 Resp: 30660
 Ion Ratio Lower Upper
 75 100
 77 0.0 0.0 0.0





#77
Bromobenzene
Concen: 18.413 ug/l
RT: 12.536 min Scan# 1774
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01



Tgt Ion: 156 Resp: 5185
Ion Ratio Lower Upper
156 100
77 190.8 93.6 280.8
158 98.4 48.3 144.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#78
n-propylbenzene
Concen: 19.208 ug/l
RT: 12.597 min Scan# 1784
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03


Tgt Ion: 91 Resp: 280785
Ion Ratio Lower Upper
91 100
120 22.4 11.3 33.8





#79
2-Chlorotoluene
Concen: 19.038 ug/l
RT: 12.682 min Scan# 1798
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01



Tgt Ion: 91 Resp: 159129
Ion Ratio Lower Upper
91 100
126 34.0 17.2 51.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#80
1,3,5-Trimethylbenzene
Concen: 19.136 ug/l
RT: 12.737 min Scan# 1807
Delta R.T. -0.006 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03


Tgt Ion:105 Resp: 187183
Ion Ratio Lower Upper
105 100
120 49.3 24.6 73.6





#81
trans-1,4-Dichloro-2-butene
Concen: 19.450 ug/l
RT: 12.304 min Scan# 1736
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01



Tgt Ion: 75 Resp: 13174
Ion Ratio Lower Upper
75 100
53 95.6 81.0 121.4
89 48.1 36.9 55.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#82
4-Chlorotoluene
Concen: 19.206 ug/l
RT: 12.779 min Scan# 1814
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03


Tgt Ion: 91 Resp: 166239
Ion Ratio Lower Upper
91 100
126 34.1 17.0 51.0





#83
 tert-Butylbenzene
 Concen: 18.691 ug/l
 RT: 12.999 min Scan# 1850
 Delta R.T. 0.000 min
 Lab File: VY021453.D
 Acq: 07 Mar 2025 12:03




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0307SBS01



Tgt Ion:119 Resp: 16310
 Ion Ratio Lower Upper
 119 100
 91 65.0 33.2 99.6
 134 27.7 13.5 40.5




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
 Supervised By :Semsettin Yesilyurt 03/11/2025



#84
 1,2,4-Trimethylbenzene
 Concen: 19.109 ug/l
 RT: 13.048 min Scan# 1858
 Delta R.T. 0.000 min
 Lab File: VY021453.D
 Acq: 07 Mar 2025 12:03

Tgt Ion:105 Resp: 185591
 Ion Ratio Lower Upper
 105 100
 120 45.7 22.7 68.1









#85
 sec-Butylbenzene
 Concen: 19.192 ug/l
 RT: 13.176 min Scan# 1879
 Delta R.T. 0.000 min
 Lab File: VY021453.D
 Acq: 07 Mar 2025 12:03

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0307SBS01

Tgt Ion:105 Resp: 24918
 Ion Ratio Lower Upper
 105 100
 134 20.0 9.9 29.7




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
 Supervised By :Semsettin Yesilyurt 03/11/2025



#86
 p-Isopropyltoluene
 Concen: 18.939 ug/l
 RT: 13.292 min Scan# 1898
 Delta R.T. -0.006 min
 Lab File: VY021453.D
 Acq: 07 Mar 2025 12:03


Tgt Ion:119 Resp: 202582
 Ion Ratio Lower Upper
 119 100
 134 25.9 13.1 39.1
 91 24.6 12.2 36.6





#87
1,3-Dichlorobenzene
Concen: 18.798 ug/l
RT: 13.292 min Scan# 1898
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01




Tgt Ion:146 Resp: 105372
Ion Ratio Lower Upper
146 100
111 40.9 20.0 60.0
148 64.4 31.8 95.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025




#88
1,4-Dichlorobenzene
Concen: 19.131 ug/l
RT: 13.371 min Scan# 1911
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03



Tgt Ion:146 Resp: 104875
Ion Ratio Lower Upper
146 100
111 41.0 19.6 58.8
148 65.4 32.3 96.8









#89
n-Butylbenzene
Concen: 18.856 ug/l
RT: 13.621 min Scan# 1952
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03

Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01

Tgt Ion: 91 Resp: 186240
Ion Ratio Lower Upper
91 100
92 52.8 26.1 78.3
134 25.4 12.6 37.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#90
Hexachloroethane
Concen: 19.087 ug/l
RT: 13.883 min Scan# 1995
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03


Tgt Ion:117 Resp: 43553
Ion Ratio Lower Upper
117 100
201 73.1 37.4 112.2





#91
 1,2-Dichlorobenzene
 Concen: 18.995 ug/l
 RT: 13.663 min Scan# 1959
 Delta R.T. 0.000 min
 Lab File: VY021453.D
 Acq: 07 Mar 2025 12:03




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0307SBS01




Tgt Ion:146 Resp: 9125
 Ion Ratio Lower Upper
 146 100
 111 41.8 20.7 62.1
 148 64.6 32.1 96.5

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 03/11/2025
 Supervised By :Semsettin Yesilyurt 03/11/2025




#92
 1,2-Dibromo-3-Chloropropane
 Concen: 19.835 ug/l
 RT: 14.273 min Scan# 2059
 Delta R.T. -0.006 min
 Lab File: VY021453.D
 Acq: 07 Mar 2025 12:03




Tgt Ion: 75 Resp: 6311
 Ion Ratio Lower Upper
 75 100
 155 77.0 43.5 130.5
 157 99.8 54.4 163.2





#93
 1,2,4-Trichlorobenzene
 Concen: 17.977 ug/l
 RT: 14.925 min Scan# 2166
 Delta R.T. 0.000 min
 Lab File: VY021453.D
 Acq: 07 Mar 2025 12:03




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0307SBS01



Tgt Ion:180 Resp: 4717
 Ion Ratio Lower Upper
 180 100
 182 96.3 46.7 140.0
 145 30.2 15.6 46.8




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 03/11/2025
 Supervised By :Semsettin Yesilyurt 03/11/2025



#94
 Hexachlorobutadiene
 Concen: 19.123 ug/l
 RT: 15.029 min Scan# 2183
 Delta R.T. 0.000 min
 Lab File: VY021453.D
 Acq: 07 Mar 2025 12:03


Tgt Ion:225 Resp: 31456
 Ion Ratio Lower Upper
 225 100
 223 61.9 31.4 94.2
 227 63.8 31.8 95.4





#95
Naphthalene
Concen: 18.887 ug/l
RT: 15.151 min Scan# 2203
Delta R.T. -0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03




Instrument :
MSVOA_Y
ClientSampleId :
VY0307SBS01




Tgt Ion:128 Resp: 74800
Ion Ratio Lower Upper
128 100
127 13.4 10.6 15.8
129 10.7 8.6 13.0

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 03/11/2025
Supervised By :Semsettin Yesilyurt 03/11/2025



#96
1,2,3-Trichlorobenzene
Concen: 18.107 ug/l
RT: 15.334 min Scan# 2233
Delta R.T. 0.000 min
Lab File: VY021453.D
Acq: 07 Mar 2025 12:03



Tgt Ion:180 Resp: 39961
Ion Ratio Lower Upper
180 100
182 95.4 47.2 141.6
145 33.3 16.4 49.1

