

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030822\
 Data File : VY007785.D
 Acq On : 08 Mar 2022 17:15
 Operator : SY/MD
 Sample : N1806-07
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DUM-25

Quant Time: Mar 09 00:28:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 01 09:08:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	3340	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.697	114	5858	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	5036	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	1076	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	2558	73.155	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	146.320%
35) Dibromofluoromethane	7.715	113	2081	55.564	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.120%
50) Toluene-d8	10.185	98	6169	44.747	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.500%
62) 4-Bromofluorobenzene	12.477	95	1402	31.549	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	63.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030822\
Data File : VY007785.D
Acq On : 08 Mar 2022 17:15
Operator : SY/MD
Sample : N1806-07
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DUM-25

Quant Time: Mar 09 00:28:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 01 09:08:03 2022
Response via : Initial Calibration

