

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030822\
 Data File : VY007790.D
 Acq On : 08 Mar 2022 19:12
 Operator : SY/MD
 Sample : N1763-04
 Misc : 5.62g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 27-28-SAMPLE-LOCATION-14

Quant Time: Mar 09 00:29:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 01 09:08:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	95472	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	142806	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	120746	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	46934	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	49612	49.637	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.280%
35) Dibromofluoromethane	7.722	113	48407	53.019	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.040%
50) Toluene-d8	10.179	98	169271	50.366	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.740%
62) 4-Bromofluorobenzene	12.483	95	44368	40.955	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	81.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030822\
Data File : VY007790.D
Acq On : 08 Mar 2022 19:12
Operator : SY/MD
Sample : N1763-04
Misc : 5.62g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
27-28-SAMPLE-LOCATION-14

Quant Time: Mar 09 00:29:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 01 09:08:03 2022
Response via : Initial Calibration

