

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030822\
 Data File : VY007795.D
 Acq On : 08 Mar 2022 21:09
 Operator : SY/MD
 Sample : N1763-10
 Misc : 5.53g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 5-6-SAMPLE-LOCATION-3

Quant Time: Mar 09 00:30:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 01 09:08:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	52658	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	79435	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	64282	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	23106	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	27947	50.695	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.380%
35) Dibromofluoromethane	7.722	113	26708	52.590	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.180%
50) Toluene-d8	10.179	98	89656	47.959	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.920%
62) 4-Bromofluorobenzene	12.477	95	21935	36.401	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	72.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030822\
Data File : VY007795.D
Acq On : 08 Mar 2022 21:09
Operator : SY/MD
Sample : N1763-10
Misc : 5.53g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
5-6-SAMPLE-LOCATION-3

Quant Time: Mar 09 00:30:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 01 09:08:03 2022
Response via : Initial Calibration

