

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030822\
 Data File : VY007797.D
 Acq On : 08 Mar 2022 21:56
 Operator : SY/MD
 Sample : N1763-12
 Misc : 5.66g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 9-10-SAMPLE-LOCATION-5

Quant Time: Mar 09 00:30:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 01 09:08:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	221613	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	366622	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	340444	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	147900	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	134278	57.876	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.760%
35) Dibromofluoromethane	7.716	113	127670	54.468	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.940%
50) Toluene-d8	10.179	98	444661	51.536	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.080%
62) 4-Bromofluorobenzene	12.483	95	128199	46.094	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030822\
Data File : VY007797.D
Acq On : 08 Mar 2022 21:56
Operator : SY/MD
Sample : N1763-12
Misc : 5.66g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
9-10-SAMPLE-LOCATION-5

Quant Time: Mar 09 00:30:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 01 09:08:03 2022
Response via : Initial Calibration

