



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030822\
Data File : VY007771.D
Acq On : 08 Mar 2022 10:46
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
Title : SW846 8260
Last Update : Tue Mar 01 09:08:03 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.9	6815	PASS
75	95	30	60	56.1	18283	PASS
95	95	100	100	100.0	32608	PASS
96	95	5	9	7.0	2269	PASS
173	174	0.00	2	1.1	358	PASS
174	95	50	100	97.4	31757	PASS
175	174	5	9	8.1	2566	PASS
176	174	95	101	96.0	30477	PASS
177	176	5	9	6.8	2060	PASS