

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030823\
 Data File : VY012884.D
 Acq On : 08 Mar 2023 20:19
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 09 00:40:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	85	0.00
2 T	Dichlorodifluoromethane	0.493	0.466	5.5	85	0.00
3 P	Chloromethane	0.589	0.540	8.3	85	0.00
4 C	Vinyl Chloride	0.540	0.538	0.4#	87	0.00
5 T	Bromomethane	0.328	0.339	-3.4	92	0.00
6 T	Chloroethane	0.324	0.331	-2.2	89	0.00
7 T	Trichlorofluoromethane	0.897	0.891	0.7	88	0.00
8 T	Diethyl Ether	0.314	0.327	-4.1	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.540	0.532	1.5	87	0.00
10 T	Methyl Iodide	0.703	0.692	1.6	80	0.00
11 T	Tert butyl alcohol	0.070	0.065	7.1	99	0.00
12 CM	1,1-Dichloroethene	0.515	0.516	-0.2#	87	0.00
13 T	Acrolein	0.087	0.077	11.5	77	0.00
14 T	Allyl chloride	0.951	0.961	-1.1	86	0.00
15 T	Acrylonitrile	0.166	0.186	-12.0	96	0.00
16 T	Acetone	0.195	0.187	4.1	76	0.00
17 T	Carbon Disulfide	1.753	1.654	5.6	82	0.00
18 T	Methyl Acetate	0.580	0.651	-12.2	95	0.00
19 T	Methyl tert-butyl Ether	1.330	1.468	-10.4	92	0.00
20 T	Methylene Chloride	0.714	0.647	9.4	91	0.00
21 T	trans-1,2-Dichloroethene	0.560	0.566	-1.1	87	0.00
22 T	Diisopropyl ether	1.727	1.830	-6.0	89	0.00
23 T	Vinyl Acetate	1.073	1.177	-9.7	90	0.00
24 P	1,1-Dichloroethane	0.984	1.012	-2.8	88	0.00
25 T	2-Butanone	0.241	0.262	-8.7	89	0.00
26 T	2,2-Dichloropropane	0.874	0.861	1.5	87	0.00
27 T	cis-1,2-Dichloroethene	0.600	0.648	-8.0	91	0.00
28 T	Bromochloromethane	0.360	0.389	-8.1	91	0.00
29 T	Tetrahydrofuran	0.146	0.169	-15.8	95	0.00
30 C	Chloroform	0.972	1.040	-7.0#	91	0.00
31 T	Cyclohexane	0.951	0.888	6.6	83	0.00
32 T	1,1,1-Trichloroethane	0.879	0.911	-3.6	89	0.00
33 S	1,2-Dichloroethane-d4	0.496	0.468	5.6	82	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	86	0.00
35 S	Dibromofluoromethane	0.302	0.281	7.0	81	0.00
36 T	1,1-Dichloropropene	0.506	0.508	-0.4	87	0.00
37 T	Ethyl Acetate	0.319	0.356	-11.6	99	0.00
38 T	Carbon Tetrachloride	0.522	0.544	-4.2	89	0.00
39 T	Methylcyclohexane	0.601	0.603	-0.3	84	0.00
40 TM	Benzene	1.430	1.485	-3.8	89	0.00
41 T	Methacrylonitrile	0.172	0.193	-12.2	95	0.01
42 TM	1,2-Dichloroethane	0.430	0.459	-6.7	91	0.00
43 T	Isopropyl Acetate	0.569	0.627	-10.2	93	0.00
44 TM	Trichloroethene	0.377	0.392	-4.0	89	0.00
45 C	1,2-Dichloropropane	0.361	0.376	-4.2#	91	0.00
46 T	Dibromomethane	0.206	0.223	-8.3	92	0.00
47 T	Bromodichloromethane	0.485	0.518	-6.8	92	0.00
48 T	Methyl methacrylate	0.262	0.303	-15.6	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030823\
 Data File : VY012884.D
 Acq On : 08 Mar 2023 20:19
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 09 00:40:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	101	0.00
50 S	Toluene-d8	1.103	0.987	10.5	79	0.00
51 T	4-Methyl-2-Pentanone	0.308	0.362	-17.5	96	0.00
52 CM	Toluene	0.863	0.915	-6.0#	90	0.00
53 T	t-1,3-Dichloropropene	0.509	0.550	-8.1	91	0.00
54 T	cis-1,3-Dichloropropene	0.576	0.610	-5.9	89	0.00
55 T	1,1,2-Trichloroethane	0.287	0.313	-9.1	94	0.00
56 T	Ethyl methacrylate	0.382	0.445	-16.5	94	0.00
57 T	1,3-Dichloropropane	0.489	0.539	-10.2	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.201	0.229	-13.9	92	0.00
59 T	2-Hexanone	0.224	0.261	-16.5	93	0.00
60 T	Dibromochloromethane	0.349	0.382	-9.5	94	0.00
61 T	1,2-Dibromoethane	0.276	0.305	-10.5	94	0.00
62 S	4-Bromofluorobenzene	0.372	0.356	4.3	83	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
64 T	Tetrachloroethene	0.364	0.368	-1.1	90	0.00
65 PM	Chlorobenzene	1.049	1.069	-1.9	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.388	0.399	-2.8	91	0.00
67 C	Ethyl Benzene	1.845	1.903	-3.1#	90	0.00
68 T	m/p-Xylenes	0.710	0.732	-3.1	89	0.00
69 T	o-Xylene	0.662	0.704	-6.3	93	0.00
70 T	Styrene	1.105	1.184	-7.1	91	0.00
71 P	Bromoform	0.253	0.276	-9.1	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.544	3.635	-2.6	90	0.00
74 T	N-amyl acetate	1.115	1.207	-8.3	94	0.00
75 P	1,1,2,2-Tetrachloroethane	0.797	0.856	-7.4	97	0.00
76 T	1,2,3-Trichloropropane	0.597	0.659	-10.4	99	0.00
77 T	Bromobenzene	0.861	0.878	-2.0	92	0.00
78 T	n-propylbenzene	4.369	4.454	-1.9	90	0.00
79 T	2-Chlorotoluene	2.428	2.488	-2.5	91	0.00
80 T	1,3,5-Trimethylbenzene	2.924	3.031	-3.7	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.288	0.299	-3.8	92	0.00
82 T	4-Chlorotoluene	2.513	2.567	-2.1	91	0.00
83 T	tert-Butylbenzene	2.524	2.647	-4.9	91	0.00
84 T	1,2,4-Trimethylbenzene	2.860	2.971	-3.9	91	0.00
85 T	sec-Butylbenzene	3.851	3.922	-1.8	90	0.00
86 T	p-Isopropyltoluene	3.147	3.221	-2.4	90	0.00
87 T	1,3-Dichlorobenzene	1.660	1.678	-1.1	91	0.00
88 T	1,4-Dichlorobenzene	1.694	1.693	0.1	91	0.00
89 T	n-Butylbenzene	3.013	3.077	-2.1	90	0.00
90 T	Hexachloroethane	0.655	0.657	-0.3	92	0.00
91 T	1,2-Dichlorobenzene	1.501	1.564	-4.2	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.129	0.146	-13.2	99	0.00
93 T	1,2,4-Trichlorobenzene	0.924	0.955	-3.4	91	0.00
94 T	Hexachlorobutadiene	0.564	0.546	3.2	90	0.00
95 T	Naphthalene	1.852	2.190	-18.3	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030823\
Data File : VY012884.D
Acq On : 08 Mar 2023 20:19
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Mar 09 00:40:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.822	0.863	-5.0	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6