

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030823\
 Data File : VY012872.D
 Acq On : 08 Mar 2023 15:39
 Operator : KP/MD
 Sample : 01829-02
 Misc : 5.66g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SS-02-20230306

Quant Time: Mar 09 00:37:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	76456	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	131131	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	120952	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	53346	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	51509	67.855	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.700%
35) Dibromofluoromethane	7.716	113	45442	57.422	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.840%
50) Toluene-d8	10.179	98	157249	54.353	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.700%
62) 4-Bromofluorobenzene	12.483	95	45737	46.838	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.680%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030823\
Data File : VY012872.D
Acq On : 08 Mar 2023 15:39
Operator : KP/MD
Sample : 01829-02
Misc : 5.66g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SS-02-20230306

Quant Time: Mar 09 00:37:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY012872.D
Acq: 08 Mar 2023 15:39




Instrument :
MSVOA_Y
ClientSampleId :
SS-02-20230306


Tgt Ion:168 Resp: 76456
Ion Ratio Lower Upper
168 100
99 57.3 41.7 62.5



#33
1,2-Dichloroethane-d4
Concen: 67.855 ug/l
RT: 8.143 min Scan# 1037
Delta R.T. -0.000 min
Lab File: VY012872.D
Acq: 08 Mar 2023 15:39

Tgt Ion: 65 Resp: 51509
Ion Ratio Lower Upper
65 100
67 51.2 0.0 104.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VY012872.D

Acq: 08 Mar 2023 15:39

Instrument :

MSVOA_Y

ClientSampleId :

SS-02-20230306





Tgt Ion:114 Resp: 131131

Ion Ratio Lower Upper

114 100

63 19.3 0.0 37.4

88 14.4 0.0 28.6



#35

Dibromofluoromethane

Concen: 57.422 ug/l

RT: 7.716 min Scan# 967

Delta R.T. 0.000 min

Lab File: VY012872.D

Acq: 08 Mar 2023 15:39




Tgt Ion:113 Resp: 45442

Ion Ratio Lower Upper

113 100

111 100.9 82.2 123.2

192 21.6 16.8 25.2





#50

Toluene-d8

Concen: 54.353 ug/l

RT: 10.179 min Scan# 1371

Delta R.T. 0.000 min

Lab File: VY012872.D

Acq: 08 Mar 2023 15:39

Instrument :

MSVOA_Y

ClientSampleId :





SS-02-20230306

Tgt Ion: 98 Resp: 157249

Ion Ratio Lower Upper

98 100

100 63.3 50.9 76.3



#62

4-Bromofluorobenzene

Concen: 46.838 ug/l

RT: 12.483 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VY012872.D

Acq: 08 Mar 2023 15:39




Tgt Ion: 95 Resp: 45737


Ion Ratio Lower Upper

95 100

174 92.6 0.0 176.6

176 88.5 0.0 170.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 117
Delta R.T. 0.000 min
Lab File: VY012872.D
Acq: 08 Mar 2023 15:39

Instrument :
MSVOA_Y
ClientSampleId :
SS-02-20230306

Tgt Ion:117 Resp: 120952
Ion Ratio Lower Upper
117 100
82 54.2 45.4 68.0
119 31.1 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY012872.D
Acq: 08 Mar 2023 15:39

Tgt Ion:152 Resp: 53346
Ion Ratio Lower Upper
152 100
115 54.8 27.1 81.3
150 156.5 0.0 347.8

