

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030922\
 Data File : VY007811.D
 Acq On : 09 Mar 2022 19:50
 Operator : SY/MD
 Sample : N1825-03
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DUM-29

Quant Time: Mar 10 01:05:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 01 09:08:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	119509	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	202769	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	184881	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	83017	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	82064	65.591	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.180%
35) Dibromofluoromethane	7.722	113	72484	55.913	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.820%
50) Toluene-d8	10.179	98	248233	52.019	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.040%
62) 4-Bromofluorobenzene	12.483	95	71402	46.418	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030922\
Data File : VY007811.D
Acq On : 09 Mar 2022 19:50
Operator : SY/MD
Sample : N1825-03
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DUM-29

Quant Time: Mar 10 01:05:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 01 09:08:03 2022
Response via : Initial Calibration

