

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031022\
 Data File : VY007829.D
 Acq On : 10 Mar 2022 16:46
 Operator : SY/MD
 Sample : N1825-03RE
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DUM-29RE

Quant Time: Mar 10 23:57:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 01 09:08:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	4562	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	6521	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	5973	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	1665	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	3965	83.019	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	166.040%#
35) Dibromofluoromethane	7.722	113	2540	60.925	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	121.840%
50) Toluene-d8	10.185	98	8162	53.184	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.360%
62) 4-Bromofluorobenzene	12.483	95	2091	42.269	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031022\
Data File : VY007829.D
Acq On : 10 Mar 2022 16:46
Operator : SY/MD
Sample : N1825-03RE
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DUM-29RE

Quant Time: Mar 10 23:57:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 01 09:08:03 2022
Response via : Initial Calibration

