

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031022\
 Data File : VY007830.D
 Acq On : 10 Mar 2022 17:09
 Operator : SY/MD
 Sample : N1829-03RE
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SU-04-030722RE

Quant Time: Mar 10 23:58:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 01 09:08:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	17242	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	26223	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	23601	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	9063	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	12382	68.595	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 137.200%	
35) Dibromofluoromethane	7.716	113	10695	63.793	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 127.580%	
50) Toluene-d8	10.179	98	31906	51.700	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 103.400%	
62) 4-Bromofluorobenzene	12.483	95	8093	40.683	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 81.360%	
Target Compounds						
95) Naphthalene	15.233	128	5273	14.248	ug/l	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031022\
Data File : VY007830.D
Acq On : 10 Mar 2022 17:09
Operator : SY/MD
Sample : N1829-03RE
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SU-04-030722RE

Quant Time: Mar 10 23:58:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 01 09:08:03 2022
Response via : Initial Calibration

