

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031022\
 Data File : VY007835.D
 Acq On : 10 Mar 2022 19:06
 Operator : SY/MD
 Sample : N1763-10
 Misc : 5.68g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 5-6-SAMPLE-LOCATION-3

Quant Time: Mar 10 23:59:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 01 09:08:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	211210	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	371028	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	337879	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	139848	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	144824	65.496	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.000%
35) Dibromofluoromethane	7.722	113	126334	53.258	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.520%
50) Toluene-d8	10.179	98	460684	52.759	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.520%
62) 4-Bromofluorobenzene	12.483	95	133348	47.376	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031022\
Data File : VY007835.D
Acq On : 10 Mar 2022 19:06
Operator : SY/MD
Sample : N1763-10
Misc : 5.68g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
5-6-SAMPLE-LOCATION-3

Quant Time: Mar 10 23:59:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 01 09:08:03 2022
Response via : Initial Calibration

