

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031122\
 Data File : VY007851.D
 Acq On : 11 Mar 2022 13:53
 Operator : SY/MD
 Sample : N1847-05
 Misc : 8.88g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CG-B1-030722-0-1

Quant Time: Mar 11 23:21:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 01 09:08:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	220640	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	377887	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	335891	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	139331	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	132246	57.252	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.500%
35) Dibromofluoromethane	7.722	113	127891	52.936	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.880%
50) Toluene-d8	10.179	98	471037	52.966	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.940%
62) 4-Bromofluorobenzene	12.483	95	128410	44.794	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031122\
Data File : VY007851.D
Acq On : 11 Mar 2022 13:53
Operator : SY/MD
Sample : N1847-05
Misc : 8.88g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CG-B1-030722-0-1

Quant Time: Mar 11 23:21:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 01 09:08:03 2022
Response via : Initial Calibration

