



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031122\
 Data File : VY007840.D
 Acq On : 11 Mar 2022 08:52
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022822S.M
 Title : SW846 8260
 Last Update : Tue Mar 01 09:08:03 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	24.7	4244	PASS
75	95	30	60	57.9	9948	PASS
95	95	100	100	100.0	17180	PASS
96	95	5	9	7.0	1206	PASS
173	174	0.00	2	1.7	275	PASS
174	95	50	100	92.1	15827	PASS
175	174	5	9	7.1	1128	PASS
176	174	95	101	98.2	15550	PASS
177	176	5	9	6.3	987	PASS